A-A-51445
12 MAY 1986

COMMERCIAL ITEM DESCRIPTION

KIT, DISPOSABLE, LENSECTOMY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable kit used in microsurgical procedures using pars plana approaches.

Salient characteristics:

General. The kit shall be a sterile, disposable, pre-assembled, microsurgical lensectomy kit compatible with an intraocular suction-cutter unit control console providing variable cutting speed of 60 to 480 cuts per minute, pressure and suction, and 20 gage intraocular cutting probe.

a. Disposable plastic bottle with plastic cap for collection of eye surgical debris and irrigating fluid having a 220 cc capacity. The bottle shall contain a self shut-off float to prevent overflow at any vacuum level. Shall provide a male Luer fitting for connection to debris line. Shall be marked from 20 cc to 180 cc in 20 cc increments.

b. Suction-cutter/fragmentation combined flexible, silicone, debris tubing in 6 foot length. Shall allow concurrent attachment of the suction-cutter probe and fragmentation handpiece with both debris lines provided with Luer fittings two feet from the probe and handpiece.

c. One disposable stiletto blade, widened to a rounded 20 gauge needle size with a plastic handle. The blade shall be fabricated of Class 3, Type 302 corrosion-resisting steel properly passivated, stress relieved and of adequate hardness to allow suitable strength and sharpness. Blade shall be suitably sharpened for its intended use.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

d. One 20 gauge needle with needle cleaning wire. Needle 20 gauge portion shall be 1 inch long with an overall length of 2 inches. Shall be threaded on the proximal end for attachment to a handpiece. Flats shall be provided on the expanded proximal section for tightening into the handpiece. Needle and wire shall be Class 3, Type 304 corrosion-resisting steel.

e. One 20 cc disposable plastic syringe with markings at each cc increment from 1 to 20 cc.

f. One three-way plastic stopcock with Luer fittings to attach to syringe and debris line fittings. Control handle shall be marked with flow direction arrows and off position.

Material. Corrosion-resisting steel shall be in accordance with ASTM Designation F899, Stainless Steel Billet, Bar, and Wire For Surgical Instruments.

Luer fittings. Luer fitting shall be in accordance with ANSI/HIMA MD70.1

Sterility. The kit components shall be sterile when tested in accordance with the United States Pharmacopeia (USP).

Safety. The needle and internal surfaces of the tubing, syringe and stopcock shall be non-pyrogenic and non-toxic in accordance with USP requirements. Plastic components coming in contact with fluids entering the eye shall contain no more than 5 parts per million of lead.

Workmanship. The kit shall be free from defects which detract from its appearance or impair its serviceability. Plastic components shall be free from dents, kinks, pits, blisters, cuts, holes and scratches. Metal components shall be free from dents, pits, burs, slivers, scale and corrosion. Needle shall not be bent, shall be sharp and shall be free from feathered edges.

Unit. Package (PG). One package containing six kits, each kit individually packaged sterile in a peel-open type container, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the currentversion before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one lensectomy kit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-166-3515	LENSECTOMY KIT, Disposable, 6s

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order .

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6515-3334

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:55Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51445	KIT, DISPOSABLE, LENSECTOMY

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assistdla.mil - Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.