GG-C-00[illegible]48a(DSA-DM)
12 September 1967
INTERIM REVISION OF
Fed. Spec. GG-C-48
27 July 1950

INTERIM FEDERAL SPECIFICATION

CABINET, TOOTH ASSORTMENT, COMPARTMENTED

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-C-48.

1. SCOPE

1.1 Scope. This specification covers a compartmented cabinet suitable for the storage of artificial tooth assortments.

2. APPLICABLE DOCUMENTS

2.1 The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-S-698	Steel, Sheet and Strip, Low-Carbon.
TT-E-491	Enamel; Gloss, Synthetic, (For Metal and Wood Hospital Furniture).
TT-P-659	Primer-Coating and Surfacer; Synthetic, Tints and White, (For Metal and Wood Surfaces).
MMM-A-250	Adhesive, Water-Resistant, (For Sealing Fiberboard Boxes).

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:48Z
Check the source to verify that this is the current version before use.

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (For Carton Sealing).

STANDARDS

Federal

FED-STD-123 Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center

24015 Cabinet, Tooth Assortment, Compartmented.

(Copies of Military standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Steel. The steel shall conform to the requirements of QQ-S-698 for cold-rolled steel.

3.1.2 Gloss enamel. The gloss enamel shall conform to the requirements of TT-E-491 for class B.

3.1.3 Primer. The primer shall conform to the requirements of TT-P-659.

3.1.4 Other material. All other material shall be of high commercial quality throughout and shall be suitable in every respect for the purpose intended.

3.2 Cabinet. The cabinet shall be constructed of steel. The cabinet shall be of welded construction, and of the design and construction shown in Defense Personnel Support Center (DPSC) drawing 24015. The cabinet shall be constructed so that two or more cabinets can be stacked without shifting horizontally.

3.3 Replacement stock control cards. Each cabinet shall be furnished with one thousand unprinted cards, 1 inch wide and 3-1/2 inches long.

3.4 Finish. The inside and outside surfaces of the cabinet shall be finished with a coat of primer surfacer of neutral tint conforming to the requirements of TT-P-659, followed by two coats of enamel of a commercial gray color, conforming to the requirements of TT-E-491. The drawer knobs and card holders shall be brass, nickel or chrome plated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

3.5 Identification marking. Each cabinet shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

3.6 Workmanship. Workmanship shall be first class throughout. The cabinets shall be free from any imperfections which detract from their appearance or which may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL) shall be as indicated in table I.

Table I. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor A	II	2.5
Minor B	S-2	4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

4.2.2 For testing. Unless otherwise specified in subsidiary specifications, testing shall be conducted as indicated for inspection level S-1 in MIL-STD-105. Any failure shall be cause for rejection.

4.3 Examination. Cabinets shall be examined for defects including, but not limited to, those listed in table II.

Table II. Classification of Defects

Categories	Defects
Major:	
101	Spot welds missing or not complete on top, bottom and side panels.
102	Cabinet framework not square and level.
103	Sharp edges, burrs, and slivers present on cabinet and drawers.
104	Drawer support channels not suitably welded to cabinet.
105	Card holders not properly attached to drawers.
106	Drawers do not move readily.
107	Cabinet dented or appearance otherwise marred.
108	Components missing.
Minor A:	
201	Plating on knob and card holder flaking or peeling.
202	Plating of knob and card holder not as specified.
203	Cabinet surfaces not free from rough edges.
204	Enamel finish not as specified.
205	Panel corners not rounded off after welding.
206	Identification marking not as specified or missing.
207	No provision for stacking.
Minor B:	
221	Component part dimensions not in agreement with gages specified.
222	Cabinet and drawer dimensions not as specified.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1.1 Unit of issue. One cabinet, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

5.1.2 Level A.

5.1.2.1 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Box shall be designed with suitable interior fittings, or shall contain sufficient cushioning material to insure against shifting of, or damage to the contents. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Each unit shall be packed in an exterior container designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Each unit shall be packed in an exterior container conforming to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with FED-STD-123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The cabinet covered by this specification is intended for use in dental prosthetic laboratories for segregating and filing assorted teeth.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:48Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

GG-C-0048a(DSA-DM)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 68 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T11:48Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

EFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

)PSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:48Z
Check the source to verify that this is the current version before use.