
FED. SUP CLASS
6640



MATERIAL: The rack shall be made of high impact polystyrene.

DESIGN: The rack is 286 x 216 x 29 mm high, outside dimensions. The center support bar holds very small pipettes, while scalloped outer edges hold 7 large and 16 smaller pipettes.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 26 JUL 1982 REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE RACK, PIPET	MILITARY STANDARD MS36391
Other Cust MS 99			
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0797
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:39Z
Check the source to verify that this is the current version before use.