FED. SUP CLASS
6640



MATERIAL: Rack shall be made of high density polyethylene plastic.

DESIGN: The rack shall have two sloping faces with drain troughs and tubes. There shall be 30 pegs on each face, three rows of eight pegs and one row of six pegs. The pegs should be approximately 6.35 mm diameter and 76 mm long. The overall dimensions are 560 mm wide x 640 mm high x 250 mm deep. The rack is used on back-to-back laboratory tables.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES:
ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES:
ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 26 JUL 1982 **REVISED**

P.A. EA **Other Cust** MS 99	**INTERNATIONAL INTEREST**	**TITLE** RACK, DRYING, LABORATORY	**MILITARY STANDARD** MS36399
Procurement Specifications NONE	**SUPERSEDES:**		**SHEET** 1 **OF** 1

DD FORM 1 MAY 73 **672** *(COORDINATED)* PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0797
STANDARD **DATE**

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:40Z
Check the source to verify that this is the current version before use.