DDD-P-001024(DSA-DM)
15 June 1965

INTERIM FEDERAL SPECIFICATION

PAD, MEDICATED

This interim Federal specification was developed by the Defense Medical Supply Center, 3rd Avenue and 29th Street, Brooklyn, New York 11232, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

1. SCOPE

1.1 This specification covers a medicated pad for use as a wet dressing on inflamed surfaces for which a bland wet dressing or cleansing pad may be indicated.

2. APPLICABLE SPECIFICATIONS, STANDARDS, AND OTHER PUBLICATIONS

2.1 Specifications and standards. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

Federal Specifications:

CCC-T-191	Textile Test Methods.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specification:

MIL-L-10547	Linens, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Military Standards:

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of Military specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN PHARMACEUTICAL ASSOCIATION

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.)

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE, FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act.

(Application for copies should be addressed to the Food and Drug Administration, Washington, D. C. 20204.)

U. S. PHARMACOPOEIAL CONVENTION, INC.

Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Flannel. The flannel shall be a plain weave outing flannel, double napped, bleached, nontoxic, absorbency treated, and shall meet the requirements specified herein.

3.1.1.1 Weight (ounces per square yard). The weight of material (see 3.1.1) shall be not less than 3.35 ounces per square yard when tested as described in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

3.1.2 Medicated solution. The solution for impregnation of the pads shall be of the following formula:

Hamamelis Water, NF		50 percent
Glycerin, USP		10 percent
Methylparaben, USP	q.s.	1:1000
Benzalkonium Chloride, USP	q.s.	1:30000
Purified Water, USP	q.s.	100 percent

The above solution shall have a pH of 4.2 to 4.8, buffered if necessary with citric acid and sodium citrate as described in the Pharmacopeia of the United States (USP), or other suitable USP or National Formulary (NF) buffer.

3.2 Dimensions. Pads shall be 3 inches $\pm$ 1/32 inch in diameter.

3.3 Construction. Each pad shall be cut from one thickness of flannel as specified in 3.1.1 and shall be discoid in shape. The pads shall be thoroughly impregnated with the medicated solution as specified in 3.1.2.

3.4 Workmanship. Workmanship shall be first class throughout. Medicated pads shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

4.1.3 Certificate of compliance. The manufacturer shall submit a certificate of compliance stating that the ingredients used in the preparation of the medicated solution meet the NF or the USP requirements as specified in 3.1.2.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

4.3 Examination. The medicated pads shall be examined to determine compliance with specification requirements.

4.4 Tests.

4.4.1 Weight of cloth. Weight of cloth may be determined in accordance with either methods 5040 or 5041 of CCC-T-191.

5. PREPARATION FOR DELIVERY

5.1 Immediate container. One hundred (100) pads shall be contained in a high impact, opaque, polystyrene jar. The jar shall have a screw on cover which may be constructed of the same material as the jar, phenol-resin, or urea. The screw on the cover shall have a resilient impervious pulp and vinyl liner. All components of the immediate container shall be compatible, and shall not interact physically or chemically with the contained substances. The strength, quality, purity, or appearance of the contents shall not be affected by the immediate container, nor shall the immediate container components be altered in any manner.

5.1.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug, and Cosmetic Act and, in addition, shall be marked with stock number, item identification, lot (control) number, date of manufacture (month, year), name and address of manufacturer, and such instructions for use as the manufacturer may deem necessary. The required information may be imprinted or applied by means of a label. All markings shall be clear, legible, and shall remain entirely affixed to the container after being immersed in a bath of water at a temperature of $75^{\circ} \pm 5^{\circ}$ for a period of not less than 4 hours. After drying, the markings shall remain entirely and permanently affixed to the containers, and the markings shall be clear and legible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

5.2 Packaging.

5.2.1 Unit of issue. One (1) jar, as specified, constitutes one unit of issue.

5.2.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I - Packaging quantities

Unit package	Intermediate package	Exterior container
1	12	24

5.2.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.3 Level A.

5.2.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

5.3 Packing.

5.3.1 Level A.

5.3.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.3.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.3.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.3.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.3.2 Level B.

5.3.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.4 Marking.

5.4.1 Interior package. Each interior package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown. Lot (control) numbers shall also be shown. In addition, the package shall be marked with arrows as specified in 5.6.7.3 of MIL-STD-129.

5.4.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) numbers shall also be shown. In addition, the container shall be marked with arrows as specified in 5.6.7.3 of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packing (see 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

Military custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.