DDD-S-001298 (VAMI)
AMENDMENT-1
July 31, 1970

AMENDMENT

TO

INTERIM FEDERAL SPECIFICATION

SPONGE, SURGICAL

This amendment was developed by the Veterans Administration, Department of Medicine and Surgery, Supply Service, Marketing Division for Medical-Dental-Scientific Supplies, Marketing Center, P.O. Box 76, Hines, Illinois 60141. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

Page 7, paragraph 3.1.5, Delete lines 9 through 13 and substitute: "per one inch width in the cross direction. The envelope shall be constructed of two pieces of paper, heat sealed on all four sides. Alternatively, the envelope may be constructed of two pieces of paper, latex sealed on three sides, and heat sealed on the fourth side at the location where the envelope is initially to be opened. Employing a thermoplastic coating, or combination of cohesive and thermoplastic coatings, a welded bond is formed, having no breaks or unsealed patterns in order to provide a secure, tamperproof, sterile maintaining closure.

Page 8, paragraph 3.1.5.1, Delete lines 5 and 6, and substitute: "averaging not less than 15 points. The envelope may be constructed of one or two pieces of paper The top and bottom of the".

480066

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:35Z
Check the source to verify that this is the current version before use.