MILITARY SPECIFICATION

SURVEYING INSTRUMENT, DISTANCE-MEASURING, ELECTRONIC MICROWAVE

This amendment forms a part of Military Specification MIL-S-52156C, dated 18 November 1976, and is approved for use by all Departments and Agencies of the Department of Defense.

PAGE 5

3.1, line 8, delete: "the control unit and"; lines 9 and 10, delete: "but still keep the instrument's zero point plumbed over the marker"; line 12, delete "not be longer than 1 meter (3 feet)" and substitute "be 5 meters (16.4 feet)"; subparagraph (a), line 5, delete: "corrected"; subparagraph (b) (1), delete in its entirety; subparagraph (b) (2), line 2, delete: "with the adapter unit indirectly mounted on a tripod"; subparagraph (c), line 3, delete "during" and substitute "between"; subparagraph (c), delete the last sentence.

PAGE 8

3.6.5, line 3, delete "7 minutes" and substitute "5 minutes".

3.7.1.1, change the maximum weight of the following items as indicated:

Item	Maximum Weight Kilograms	(Pounds)
"Antenna cables (short and long)	20	(44.1)
Control unit, antenna and adapter unit without main power source	22	(48.5)"

PAGE 9

3.7.1.2 (b), line 2, delete ".018 cubic meters (.64 cubic feet)" and substitute "0.31 cubic meters (1.11 cubic feet)"; line 3, delete "300 mm (11.8 inches)" and substitute "400 mm (15.8 inches)".

3.7.3, line 2, delete: "corrected".

3.7.4, delete the last two sentences.

Source: https://assist.dla.mil -- Downloaded: 202610-22T21:08Z
Check the source to verify that this is the current version before use.

PAGE 10

3.8.5, delete the first sentence. At the end of the paragraph, add: "unless the internal lighting has a light bulb redundency of three light bulbs for the illumination of all operating indicators and dials."

3.8.8, at the end of the paragraph, add: "The test meter shall be located on front of the control unit."

After 3.8.8, add:

"3.8.9 Signal strength meter. A signal strength meter, indicating signal return strength, used for achieving maximum signal return shall be located on the back of the antenna unit. It shall be built in accordance with MIL-M-10304. The internal illumination shall be in accordance with MIL-L-25467."

PAGE 11

3.10, at the end of the paragraph, add: "unless the power cable is wired internally to the control unit."

3.11, line 2, after "H-91/U" add: "or H-144/U,".

PAGE 12

3.12.1, delete the last two sentences.

3.12.2, line 9, delete: "control unit via adapter unit and on a".

PAGE 13

3.16, line 4, delete: "and still keep the instrument's zero point plumbed over the marker."

PAGE 21

4.6.2.2, Table II. Test Schedule, place an "X" in Column 1 for "Reliability" and for "Maintainability".

Custodians:
Army - ME
Air Force - 99

Review activities:
Army - EL, CE
DLA - GS

User activity:
Navy - MC

Preparing activity:
Army - ME

Project 6675-0645

Source: https://assist.dla.mil -- Downloaded: 202640-22T21:08Z
Check the source to verify that this is the current version before use.