A-A-51480A
30 May 1989
SUPERSEDING
A-A-51480
20 June 1986

COMMERCIAL ITEM DESCRIPTION

BAG, URINE COLLECTION, PEDIATRIC

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-B-36444 which is cancelled.

This Commercial Item Description covers a disposable plastic bag, suitable for collecting urine specimens from newborn infants and small children.

Salient characteristics:

The bag shall be lightweight, disposable, durable, clear, leakproof, non-toxic, and polyethylene, suitable for collecting urine specimens from newborn infants and small children. The bags shall be furnished in the following types:

Type I - 30 ml, with drainage tube for 24-hour sample (continuous collection).

Type II - 100 ml, for single sample.

Type III- 60 ml, with drainage tube for 24-hour sample (continuous collection).

Type IV - 200 ml, for single sample.

Note: The milliliter (ml) volumes shown above are the practical capacities.

The polyethylene film shall have a minimum tensile strength of 1500 PSI and a minimum elongation of 200 percent when tested using test method ASTM-D882. The film shall be sufficiently transparent to permit viewing of the contents of the bag.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

The bag shall be double chambered. The inner bag shall be fabricated from polyethylene film a minimum of 0.002 inch thick. The inner bag shall have 2 slits in the polyethylene film near the bottom edge. The slits shall act as check valves until the bag is filled to its practical capacity, allowing the urine to pass from the inner bag to the outer bag and discouraging the urine from re-entering the inner bag and splashing against the skin.

The Type II and Type IV outer bags shall have a 1/4 inch diameter drain hole near one bottom corner. The drain hole shall have a removable polyethylene, or polyester, film cover, at least 0.002 inch thick, which shall be securely sealed to the outer bag.

The Type I and Type III outer bags shall have a drainage tube connected to one bottom corner. The drainage tubing shall consist of the tubing, a plastic or corrosion-resisting metal junction, and a tube cap. The tubing shall be clear flexible vinyl tubing with an approximate outside diameter of 0.2 inches. One end of the tubing shall be connected to a bottom corner of the outer bag. The junction shall be capable of joining two pieces of tubing and shall be stored on the free end of the tubing when not in use. The junction may be affixed to the packaging in such a way that it cannot be misplaced upon opening the protective packaging. The cap shall be capable of stopping both the free end of the tubing and the end of the junction. The drainage tubing shall be 60 plus or minus 1 inch in length.

Type I, Type II, Type III, and Type IV urine collection bags shall have practical capacities of 30 ml, 100 ml, 60 ml, and 200 ml, respectively. Type I and Type II bags shall be 3-1/2 inches plus or minus 1/2 inch wide by 6-1/2 inches plus or minus 1/2 inch long. Type III and Type IV bags shall be 3-1/2 inches plus or minus 1/2 inch wide by 8-1/2 inches plus or minus 1/2 inch long.

The bag shall have an adhesive area surrounding the opening which shall hold the bag in position on a male or female newborn infant or small child and shall prevent fecal contamination of the urine specimen. After a specimen has been taken, it shall be possible to fold the adhesive area on itself to make a leakproof seal for transportation.

The surgical adhesive shall be pressure-sensitive, hypo-allergenic, non-toxic, and non-irritating. Upon removal of the bag from the patient, the adhesive shall not leave any residue on the surface of the skin, nor shall any abrasion or irritation of the skin result.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.

The release paper shall be a protective cover for the adhesive. The paper shall adhere firmly to the adhesive, but shall not remove any of the adhesive when it is peeled away. The paper shall not modify the quality of the surgical adhesive.

All seams in the polyethylene film except for the drainage tube connection shall be completely heat sealed.

The urine collection bag shall not leak when filled to its practical capacity with tap water and suspended on a suitable support for 24 hours. The bag shall also be free of leaks when it is filled with tap water to its practical capacity with the adhesive area folded over on itself and held in the inverted position (seal downward) for a period of 5 minutes.

The urine collection bag shall be free of holes and tears and shall not leak. The bag shall be free of oil, grease, dirt, or other contamination.

Each unit of issue (50 or 100 bags) of urine collection bags shall be supplied with a minimum of two comprehensive instruction sheets for use. As an alternate the instructions for use shall be permanently and legibly marked on each bag.

Workmanship. The pediatric urine collection bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG) Type I and III. One package containing fifty Type I or Type III bags individually sealed in a clear plastic envelope and further packaged in a dispenser box constitutes one unit.

Package (PG). Type II and IV. One package containing one hundred bags of one type only packaged in a dispenser box constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one pediatric urine collection bag, as specified in this description.

Defects.

Bag is not complete with components.

Heat seals are not complete.

Bag is not free of holes or tears.

Adhesive area is not complete.

Adhesive does not remain on bag upon removal of paper cover.

Bag is not free of oil, grease, dirt, or other contamination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one pediatric urine collection bag.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests.

Leakage. The urine collection bag shall be filled with a measured amount of tap water equal to the practical capacity of the bag. The bag shall then be suspended or a suitable support for 24 hours. At the end of this period the bag shall be examined for evidence of leakage. Any evidence of leakage shall constitute a defect.

Watertightness of adhesive seal. The urine collection bag shall be filled to its practical capacity with tap water. The backing shall then be removed from the adhesive area and the adhesive shall then be pressed to a piece of glass which has previously been cleaned with alcohol. The bag shall then be removed from the glass and the adhesive area shall be folded over on itself to form a water tight seal. The bag shall then be suspended in the inverted position (seal downward) for a period of 5 minutes. Any evidence of leakage shall constitute a defect.

Tensile strength. The polyethylene film material shall be tested for tensile strength according to test method ASTM-D882. A tensile strength less than 1500 PSI shall constitute a defect.

Elongation. The polyethylene film material shall be tested for elongation according to test method ASTM-D882. An elongation less than 200 percent shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-00-181-8099 (Type I)
6530-00-495-7143 (Type II)
6530-00-495-7144 (Type III)
6530-00-935-4120 (Type IV)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2056

Location:
CID/BAG/URINE/COLLECTION/H13-19

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 604-833/D062

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)* and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51480A	BAG, URINE COLLECTION, PEDIATRIC

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.