A-A-53413
24 March 1988

COMMERCIAL ITEM DESCRIPTION

HOLDER, ANAEROBIC JAR AND STRIP RACK

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers plastic holders for horizontally supporting six (6) strips of inoculated media in an anaerobic jar while the strips are being incubated.

Salient characteristics:

Shall be an anaerobic jar stand (holder), and a strip rack holder, that are designed to horizontally hold six (6) strips of inoculated media in an anaerobic jar while the strips are being incubated. The strip rack holder shall also provide a central location for a moistened envelope containing carbon dioxide and hydrogen tablets for the generation of the anaerobic atmosphere within the anaerobic jar.

The anaerobic jar stand and the strip rack holder shall be made of a transparent plastic material.

The anaerobic jar stand shall be designed to provide steady support for an anaerobic jar lying on its side in a horizontal position. The stand shall be dimensioned to accommodate an anaerobic jar stocklisted under (JAR, ANAEROBIC CULTURE APPARATUS), nominally 5-1/8 inches (13.0 cm) in diameter at its center, and 9 inches (22.9 cm) in overall height. (See notes)

The strip rack shall be designed for insertion into the anaerobic jar with a tab provided to facilitate insertion and removal. It shall have six (6) compartments, supported by two (2) inner rings, for holding the test strips. The inner rings shall form a centrally located compartment and shall be slotted to hold a moistened envelope containing the anaerobic atmosphere generating tablet. The dimensions of the strip holder shall be able to accommodate the following:

1. Test strips with nominal dimensions of 8-3/4 by 1-1/4 inches (22.2 by 3.2 cm), Analytab Products Inc. Part No. 8886-036253 API 20A System, or equal, stocklisted under (TEST KIT, ANAEROBE IDENTIFICATION). (See notes)

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

2. An envelope with anaerobic atmosphere generating tablets, that has nominal dimensions of 3-1/2 by 5-7/8 inches (8.9 by 14.7 cm), and is stocklisted under (GENERATOR, CARBON DIOXIDE-HYDROGEN, ANAEROBIC CULTURE APPARATUS). (See notes)

Workmanship. The anaerobic jar stand (holder) and the strip rack holder shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One anaerobic jar holder and a strip rack holder, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one anaerobic jar stand and strip rack holder.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State. and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTES: The following National Stock Number is covered by this document:

6640-01-139-7481	Holder, Anaerobic Jar and Strip Rack

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

The following National Stock Numbers shall be compatible for use with this National Stock Number:

6640-00-926-7655	Jar, Anaerobic Culture Apparatus
6640-00-935-6863	Generator, Carbon Dioxide-Hydrogen, Anaerobic Culture Apparatus
6550-01-139-7505	Test Kit, Anaerobic Identification

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1455

Location: Disk # 10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.