A-A-0053420(DM)
15 March 1988

INTERIM COMMERCIAL ITEM DESCRIPTION

CALIBRATION AND VERIFICATION SYSTEM

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a portable, battery operated, calibration and verification system for measuring to BRH standards the diagnostic X-Ray systems output. Consists of a non-invasive digital KVP meter, non-invasive exposure duration timer, and a generator milliampere second meter.

Salient characteristics:

The unit shall be a calibration/verification system to accurately measure, to BRH standards, the output of diagnostic x-ray systems. The calibration verification system shall measure kilovolt-peak (kVp), exposure duration in milisecond (ms) and/or second (s), and x-ray generator millampere-second (mAs). The calibration/verification system shall accomplish the required kilovolt-peak and exposure duration measurements non-invasively. The milliampere-second measurement may be made invasive or non-invasive. The system shall include the instruments and respective connecting cables as follows:

a. Non-invasive kVp measuring device.
b. Non-invasive exposure duration timer.
c. Generator millampere-second measuring device.

Components included in the system shall consist of a maximum of three solid state devices and shall be packed in a single protective container. The outside dimensions of the container shall be 12-18" (L) X 10-16" (W) X 4-8" (H) and have a total system weight between 5 and 18 pounds.

The system shall have a service life, to include storage, of at least seven years (excludes batteries).

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

The calibration and verification system shall have the following design features:

a. Subassemblies and chassis components which are removable without removing any other hard-wire subassembly or printed circuit board or component.

b. Permit adjustments without removing any component, printed circuit board, or module. The use of a circuit board/module extender is acceptable.

c. Printed circuit boards which are removable without the need to unsolder cables or wires.

The calibration and verification system shall have an operating temperature range of at least 15°C to 35°C, relative humidity to 90 percent. Storage temperatures shall be at least -35°C to 50°C, relative humidity to 90 percent.

The system shall be battery operated and utilize batteries that are currently available through Defense Supply System.

The system shall be capable of maintaining tolerance specifications for a period no less than 360 days.

The system shall be used in temporary or mobile shelters normally issued to medical TOE units, shall be packaged to withstand ground mobile transport over secondary and tributary roads and cross country terrain without adverse effects, and shall be compatible without additional packaging.

The High Voltage Measuring Device shall have operating capabilities as follows:

a. Measure kVp of an x-ray system without physically making electrical connections to the x-ray system in any manner.
b. Provide a signal(s) which can be displayed on an oscilloscope or digital multimeter.
c. Count and display pulse duration or number of x-ray pulses or provide an output signal for use with a timing device.
d. Be operational with single phase, three phase, half wave, full wave or self rectified x-ray system, and high frequency generators.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:19Z
Check the source to verify that this is the current version before use.

e. Frequency range of detection of the kVp meter shall range from direct current (DC), as produced by constant potential generators, to an upper limit 50 KHz, as produced by high frequency generators.
f. Shall have a range of 50-150 kVp, ± 2 percent of reading. Operating temperature range shall be 15° -35°C. Storage temperature range shall be -35°-50°c. Shall have an operating humidity range of 0-90 percent.

The time measuring device shall measure x-ray exposure times and shall have the following capabilities:

a. Pulse count range(s) of at least 0 to 9999.
b. Exposure time range(s) of at least 0 to 9.999s ±0.1 percent +1 digit.

The Milliampere Second (mAs) Measuring Device shall measure the amount of current present in the secondary of the high tension transformer of the x-ray system during an exposure period.

The milliampere-second device shall have a current range of 0-200 mAs, ± 1 percent and a current-time range of 0-2000 mAs, ± 1 percent.

Workmanship. The calibration and verification system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One calibration and verification system, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance iwht MIL-STD-105 and as specified herein. The unit of product for tests purposed shall be one calibration and verification system.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6625-01-268-0383

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6625-0694

Location: ICID/CALIBRATION/H17-31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.