A-A-53423
23 March 1988

COMMERCIAL ITEM DESCRIPTION

BAG, BLOOD WARMING (Disposable)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable blood warming bag suitable for use in a dry heat blood warmer.

Salient characteristics:

General. The blood warming bag shall consist of a plastic bag, integral tubing and a connector on the inlet side for connection to a recipient set and an outlet chamber which acts as a bubble trap with integral tubing for connection to a recipient.

Material. The bag shall be constructed of 6 mil vinyl. The tubing shall be fabricated from polyvinylchloride (PVC) plastic. The blood warming bag and integral connected items involved in direct contact with the flow of blood shall meet the requirements of the Pharmacopeia of the United States (USP) for Class IV plastics.

Style, design and dimensions. The bag and auxiliary connected items shall be compatible with a dry heat blood warmer. The bag shall be 8 ± .2 inches wide by 9.4 ± .2 inches long and shall contain a continuous convoluted channel 140 ± 2 inches long. When used with the dry heat blood warmer, the bag and tubing shall be capable of allowing the blood to be warmed to a 32° to 37° C range at flow rates up to 150 ml per minute.

The inlet tubing shall be 0.12 ± .01 inch internal diameter and 12 ± .2 inches long. The end of the tubing shall provide a plastic female Luer fitting for connection to a recipient set.

The outlet tubing shall contain a bubble trap, at the outlet of the bag, connected to 0.12 ± 0.01 inch internal diameter PVC tubing 48 ± 1 inches long. The recipient end of the tubing shall be connected to a rubber flash bulb attached to a plastic male Luer fitting. A roller-type plastic clamp shall be provided on the outlet tubing.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Sterility, pyrogenicity and toxicity. The internal surfaces of the bag, connected tubing, and fitting shall be sterile, non-pyrogenic and non-toxic when tested in accordance with the USP for transfusion and infusion assemblies.

Workmanship. The blood warming bag shall be free from defects which detract from its appearance or impair its serviceability. The blood warming bag shall be free from cuts, tears, holes, weak sections, inclusions, and bubbles. The blood warming bag shall be free from grease, dirt and foreign material.

Unit. Package (PG). One package containing twenty-four individually packaged blood warming bags, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one blood warming bag.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Note: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-139-3625	BAG, BLOOD WARMING, Disposable, 24's, Sterile

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51858

Project No. 6515-3510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.