A-A-53425
25 March 1988

COMMERCIAL ITEM DESCRIPTION

JAR, ANAEROBIC CULTURE APPARATUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a jar which is part of a self-contained anaerobic apparatus and is used to contain the anaerobic atmosphere.

Salient characteristics:

Shall be the jar component of (or jar replacement for) a self-contained anaerobic apparatus compatible for use with the following components entering into the overall apparatus: (See notes)

Lid Assembly
Carbon Dioxide-Hydrogen Generator
Indicator, Anaerobic

The jar shall be fabricated of transparent polycarbonate plastic.

The jar shall be round with an integral flange around the full circumference at the open end. The flange and wall thickness dimensions shall be adequate to safely withstand internal pressures or vacuum generated within the jar during use.

Nominal dimensions of the jar shall be 9 inches (22.9 cm) in overall height, 6 inches (15.2 cm) outside diameter of the flange, and 5-1/8 inches (13.0 cm) outside diameter below the flange.

The top shall have a smooth, flat contact surface so as to make a leakproof seal when mated with the "O" ring gasket of the lid assembly, and secured into operating position on the jar.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Workmanship. The jar shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One jar, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one jar.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-00-926-7655

The jar shall be compatible for use with the following National Stock Numbers entering into the overall anaerobic apparatus:

6640-00-001-6670	Lid Assembly
6640-00-935-6863	Carbon Dioxide-Hydrogen Generator
6640-00-145-1150	Indicator, Anaerobic

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1457

Location: Disk # 10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.