A-A-53428
31 March 1988

COMMERCIAL ITEM DESCRIPTION

CONTROL STRIP, REAGENT SUITABILITY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a quality control reagent test strip with positive and negative controls to determine the suitability of six reagents in an identification system for Enterobacteriaceae and other gram-negative bacteria.

Salient characteristics:

Shall be a quality control reagent test strip with positive and negative controls, to determine the suitability of six reagents. Shall be suitable for use with an Enterobacteriaceae Differentiation Test Kit(i.e.: Analytab API 20E).

Shall be used with the following reagents (see Notes):

(6% Alpha-Napthol reagent)
(N, N-Dimethyl-a-Naphthylamine reagent)
(Sulphanilic acid reagent)
(Potassium hydroxide reagent)
(Ferric chloride solution)
(Kovacs' reagent)

Shall be suitable for performing the following four biochemical tests:

1. Tryptophane deaminase (TDA). To check the ferric chloride solution.
2. Indole (IND). To check the Kovacs' reagent.
3. Nitrate reduction (NIT). To check the N, N-Dimethyl-a-Napthylamine and Sulphanilic acid reagents.
4. Voges-Proskauer (VP). To check the Alpha-Naphtol and Potassium hydroxide solutions.

Shall consists of eight microtubes containing dehydrated media that shall be reconstructed by adding a bacterial suspension. Four of the microtubes shall serve as positive controls for biochemical test, and the other four shall serve as negative controls for the biochemical tests.

AMSC N/A FSC 6550

DISTRIBUTION STATEMENT.A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Shall be supplied with instructions for proper use.

Workmanship. The control strips shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing thirty control strips, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one unit.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for one or more tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Biomedical test. A biomedical test shall be conducted on the control strip using the TDA, IND, NIT and VP testing methods. Sufficient quantities of the six reagents shall be used in the testing. The control strip shall be inoculated with viable cultures that have been previously identified. The testing shall be conducted in accordance with the manufacturer's instructions. The biomedical reactions on the control strip shall be recorded and compared with a suitable reference that indicates expected biochemical reactions.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6550-01-168-2605	CONTROL STRIPS, REAGENT SUITABILITY

The following National Stock Numbers are the reagents that subject NSN is intended for use with:

6550-01-139-7484	6% Alpha-Napthol
6550-01-139-7485	N, N-Dimethyl-a-Naphthylamine
6550-01-139-9544	Sulphanilic acid
6550-01-153-6987	Potassium hydroxide
6550-01-153-6988	Ferric chloride
6550-01-153-6991	Kovacs

Ordering data. (Expiration dating period required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6550-0068

Location: Disk # 10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.