A-A-53446
29 April 1988

COMMERCIAL ITEM DESCRIPTION

SYRINGE, FOUNTAIN

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Interim Federal Specification GG-S-001015a (DSA-DM) which is cancelled.

This Commercial Item Description covers a completely assembled, disposable, enema fountain syringe.

Salient characteristics:

The fountain syringe shall be a disposable enema unit, bag or bucket type, with tubing, soap, underpad and lubricant, suitable for instilling fluids rectally.

The bag type enema unit shall hold a minimum of 1200 ml of fluid; the bucket type enema unit shall hold a minimum of 1400 ml of fluid. The bag or bucket unit shall be leakproof and shall include a filling spout and outlet. The filling spout at the top shall have a suitable leakproof mechanical or self-sealing type closure.

The enema unit shall have graduation markings at not less than 3 points. Markings shall be accurate to ± 25 ml. The unit shall be supplied with a suitable, secure method to permit suspension of the filled unit from an intravenous stand or other support.

The enema unit outlet tubing shall be permanently attached at one end to the bottom opening of the bag or bucket, and there shall be no constriction of the opening as the result of the attachment of the tubing. The tubing shall be not less than 60 inches (152.4cm) long, shall be approximately 7-8 millimeters in outside diameter, and shall be marked approximately 5 inches (12.7cm) from the distal end to indicate tube depth of penetration. The mark shall not wash off with soap, water, or washcloth. There shall be not less than one opening approximately 1/8 inch in diameter, approximately 1/2 inch from the distal end of the tube. The tubing shall be flexible, non-toxic, and non-kinking.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

The enema unit shall be provided with a positive acting shut-off clamp, attached to the tubing, which by compression of the tubing shall be capable of accomplishing complete shut-off. The shut-off clamp shall be made of a suitable corrosion-resistant metal or a suitable plastic, such as polypropyene or polyethyene, of sufficient thickness that the tubing will not be cut when the clamp has closed off the path.

Two-thirds of a fluid ounce of liquid castile soap shall be supplied in each enema unit.

The enema unit shall contain one disposable, absorbent underpad, approximately 17 inches by 19 inches, folded to approximately
5 inches by 8-1/2 inches. The underpad shall have a three-ply thickness of wet strength facial tissue especially embossed to give softness and high absorbency. The pad shall be backed with a layer of soft pliable plastic film providing complete moisture barrier protection.

The enema unit shall contain a sterile lubricant which shall be supplied in a vial, a foil package, or a plastic package.

Each enema unit shall contain one set of instructions for use.

Workmanship. The enema unit fountain syringe shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One syringe, with instructions for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one fountain syringe as specified in this description.

Defects.

Any component missing or incomplete.

Sealing device at filler spout does not function properly.

Enema unit has no suitable means for suspension.

Tubing not permanently attached to bag.

Tubing not reasonably round; not free from obstruction throughout its length.

Tubing does not have smooth rounded tip integral.

Lubricant missing or not of proper consistency.

Opening in tip not in proper place, proper diameter or missing.

Graduations missing or illegible.

Instructions for use not included.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-3. The unit of product for test purposes shall be one fountain syringe.

Tests.

Clamp Test. A sample enema unit, with shut-off clamp closed, shall be filled to its nominal capacity with water at 120°F by its suspension device for 60 minutes. The unit shall then be examined for evidence of leakage and suspension failure. Any leakage or suspension failure shall constitute a defect.

Drop Test. A sample enema unit, with shut-off clamp and filling spout closed, shall be filled to its nominal capacity with warm water at 120°F and dropped from a height of two feet onto a hard regular surface. Any rupture or leak shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-073-6264

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1970

PGC 49947

Location: CID/SYRINGE/H13-25

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53446	SYRINGE, FOUNTAIN

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.