A-A-53449
29 April 1988

COMMERCIAL ITEM DESCRIPTION

TRAY, STERILIZATION, ARTHROSCOPIC INSTRUMENT COMPONENTS

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a corrosion-resisting steel tray with cover used for handling, carrying and sterilization of arthroscopic instrument components.

Salient characteristics:

The tray shall be a lightweight, stainless steel sterilization case, approximately 16 inches long by 10 inches wide by 1-5/8 inches thick (40 cm by 25.4 cm by 4.13 cm), specifically designed to hold all components of the arthroscopic intra-articular surgery system during handling and sterilization,.

The tray shall have a detachable cover permitting good visibility and easy placement on sterile surgical fields. The tray shall rest evenly on a flat surface without rock or wobble.

The top and bottom of the tray shall be perforated with hundreds of minute holes for maximum steam or ethylene oxide penetration of case contents.

The tray shall have multiple interior compartments individually shaped to the contour and sizes of the intra-articular power system components.

The tray bottom shall have a minimum of four integral V-notched metal strips welded to the tray which automatically center the instruments or scopes. The V-notched metal strips shall accommodate at least 12 instruments or scopes. Springs in the cover shall totally immoblize them when the cover is closed.

The tray shall have dual spring locks for securing the lid to the tray bottom during handling and sterilization.

The tray shall have three handles, one on each end and one centered on the front lateral panel.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

The bottom and top surfaces of the tray shall be flat, facilitating stacking during storage and use.

The tray shall be fabricated of corrosion-resisting steel, American Iron and Steel Institute (AISI) 300 series, properly hardened and passivated.

The tray shall have a #4 commercial finish or better.

All welds shall be free of cracks, pits, porosity, and other defects. All welds shall be ground smooth.

The tray shall be free from cracks, burrs, pits, dents, rough areas, scale, corrosion, and sharp edges. The finished tray shall be free from grease, oil, dirt, chips, or other foreign matter.

Note: Ther sterilization tray is to be used as a case to separate and protect all components of the arthroscopic intra-articular surgery system during handling and sterilization. In order to maintain the sterility required for surgical use, the tray shall be wrapped in some type of wrapper, envelope, or other standard method for sterilization, to allow maintenance of sterility and to be able to present tray with components on a sterile field.

Workmanship. The tray shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One sterilization tray, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one sterilization tray as specified in this description.

Defects.

Tray not manufactured of corrosion resisting steel.

Tray does not have a detachable cover.

Tray not free of cracks, tears, holes, or ripples.

Tray edges not free of burrs or sharpness.

Tray rocks, wobbles, or does not rest evenly on a flat surface.

Finish not a #4 commercial finish or better.

Finish not free of scratches or mars.

Tray not free of grease, oil, dirt, chips, or other foreign matter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-3. The unit of product for test purposes shall be one sterilization tray.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-169-5926

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1876

PGC 51047

Location: CID/TRAY/H13-24

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvement. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53449	TRAY, STERILIZATION, ARTHROSCOPIC INSTRUMENT COMPONENTS

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*. _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.