COMMERCIAL ITEM DESCRIPTION

WRAPPER, STERILIZATION

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable wrapper suitable for wrapping surgical supplies for sterilization by autoclaving.

Salient characteristics:

The sterilization wrapper shall be a clean, disposable, polypropylene, regular weight wrapper, fabricated of a non-woven material, suitable for wrapping and processing surgical supplies for sterilization. The wrapper shall be supplied in the following approximate sizes:

Size 1 - 18 inches (45.7 cm) square

Size 2 - 24 inches (61.0 cm) square

Size 3 - 36 inches (91.4 cm) square

Size 4 - 54 inches (137.2 cm) square

The wrapper shall be fluid repellent, memory-free, low linting, and puncture and abrasion resistant. The wrapper shall have bacterial barrier properties. The weight of the wrapper shall be approximately 1.3 ounces per square yard.

The wrapper shall be freely permeable to steam and ethylene oxide penetration. The wrapper shall assure efficient concentration of steam to guarantee sterilization of the contents when used as a double wrapper. The wrapper shall insure adequate air removal for drying efficacy.

The wrapper shall not have more than a three percent weight gain during steam sterilization.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

The wrapper shall maintain the supplies in a sterile manner for 30 days without a "dust cover" in a clean, temperature controlled, humidity controlled environment if the item has not been opened or tampered with.

The wrapper shall meet Electrical Surface Conductivity Standard, NFPA 99, and Flame Resistance Standard, NFPA 702, class I. The wrapper shall be nonallergenic as evidenced by no primary irritation or sensitization as evaluated by the U.S. Public Health Service, Dermatosis Section.

The wrapper shall be suitable for use with a standard sterilization indicator tape. The wrapper shall maintain its resiliency after exposure to the high temperature of sterilization, not becoming hard or brittle. It shall be resistant to tearing or puncturing after sterilization, and shall be easy to open safely while observing aseptic technique.

The wrapper shall be free of tears, cuts, holes, perforations, grease, dirt, or foreign matter.

Workmanship. The sterilization wrapper shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one thousand wrappers of size 1 only, as specified, constitutes one unit.

Unit. Package (PG). One package containing five hundred wrappers of size 2 only, as specified, constitutes one unit.

Unit. Package (PG). One package containing three hundred wrappers of size 3 only, as specified, constitutes one unit.

Unit. Package (PG). One package containing one hundred wrappers of size 4 only, as specified, constitutes one unit. The wrappers shall maintain sterility for 30 days without a "dust cover" in a clean, temperature controlled, humidity controlled environment if the item has not been opened or tampered with.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one sterilization wrapper as specified in this description.

Defects.

Wrapper not made of polypropylene.

Wrapper not free of tears, cuts, holes, or perforations.

Weight of wrapper not correct.

Wrapper not free of grease, dirt, or foreign material

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:19Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-1. The unit of product for test purposes shall be one sterilization wrapper.

Tests

Sterilization test

Typical surgical supplies shall be placed inside the sterilization wrapper, sealed with sterilizer indicator tape, and autoclaved with steam at 250°F. and 15 pounds pressure for 30 minutes. The wrapper shall then be inspected for hardness, brittleness, and resistance to puncturing or tearing. Any excessive hardness, brittleness, or tendency to puncture or tear, shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-049-1829 (Size 1)
6530-01-029-8764 (Size 2)
6530-01-049-1830 (Size 3)
6530-01-086-2464 (Size 4)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1975

PGC 05942

Location: CID/WRAPPER/H13-28

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53450	WRAPPER, STERILIZATION

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.