A-A-53451
29 April 1988

COMMERCIAL ITEM DESCRIPTION

WRAPPER, STERILIZATION

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable surgical paper wrapper suitable for wrapping surgical supplies for sterilization by autoclaving.

Salient characteristics:

The sterilization wrapper shall be made of one piece of single thickness paper, creped in one direction, suitable for wrapping surgical supplies for sterilization by autoclaving. The wrapper shall be made in the following approximate sizes:

Size 1 - 15 inches (38.1 cm) square

Size 2 - 20 inches (50.8 cm) square

Size 3 - 25 inches (63.5 cm) square

Size 4 - 40 inches (101.6 cm) square

The wrapper shall be freely permeable to steam penetration during autoclaving, and fully protect the sterilized supplies during reasonable periods of storage in non-sterile surroundings.

The wrapper shall be low linting, puncture and abrasion resistant, and have bacterial barrier properties.

The wrapper shall maintain the autoclaved supplies in a sterile manner for 30 days without a "dust cover" in a clean, temperature controlled, humidity controlled environment if the item has not been opened or tampered with.

The wrapper shall meet flame resistance standards of NFPA 702, class 3. The wrapper shall be nonallergenic as evidenced by no primary irritation or sensitization as evaluated by the U.S. Public Health Service, Dermatosis Section.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The wrapper shall be suitable for use with a standard sterilization indicator tape. The wrapper shall maintain its resiliency after exposure to the high temperature of sterilization, and not become hard or brittle. It shall be resistant to tearing or puncturing after sterilization, and shall be easy to open safely while observing aseptic technique.

The wrapper shall be free of tears, cuts, holes, perforations, grease, dirt, or foreign matter.

Workmanship. The sterilization wrapper shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 500 wrappers for NSN 6530-00-330-4507 and 6530-00-926-2234, as specified, constitutes one unit.

Package (PG). One package containing 1000 wrappers for NSN 6530-01-071-8519,, as specified, constitutes one unit.

Package (PG). One package containing 250 wrappers for NSN 6530-01-071-8520, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one sterilization wrapper as specified in this description.

Defects.

Wrapper not creped in one direction.

Wrapper not free of tears, cuts, holes, or perforations.

Wrapper not low linting.

Wrapper not free of grease, dirt, or foreign material

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-3. The unit of product for test purposes shall be one sterilization wrapper.

Tests.

Sterilization test

Typical surgical supplies shall be placed inside the sterilization wrapper, sealed with sterilization indicator tape, and autoclaved with steam at 250°F and 15 pounds pressure for 30 minutes. The wrapper shall then be inspected for hardness, brittleness, and resistance to puncturing or tearing. Any excessive hardness, brittleness, or tendency to puncture or tear, shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-071-8519 (Size 1)
6530-00-330-4507 (Size 2)
6530-00-926-2234 (Size 3)
6530-01-071-8520 (Size 4)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1977

PGC 05907

Location: CID/WRAPPER/H13-29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4906 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53451	WRAPPER, STERILIZATION

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.