A-A-53459
15 December 1988

COMMERCIAL ITEM DESCRIPTION

SUCTION APPARATUS, SURGICAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a suction apparatus suitable for use in an ambulance, hospital, or in a stationary field hospital to aspirate secretions, blood, or vomitus as necessary.

Salient characteristics:

The apparatus shall be capable of being powered via manual operation (component hand operated pump), hospital wall vacuum, or vehicle vacuum system as an oral, nasal, and tracheal suction device.

The apparatus shall be suitable for use in aspirating secretions, blood, and vomitus during emergency treatment including treatment of apnea patients.

In addition to requirements specified herein, all components, attachments, materials, and functions shall be in compliance with ASTM F960 (Standard Specification for Medical and Surgical Suction and Drainage Systems) as applicable for specified use.

The apparatus shall consist of, at least, a double cycle hand pump, a collection/vacuum container with a container liner installed, vacuum gauge, a vacuum adjustment control, a patient suction tube, a vehicle vacuum hose, all necessary interconnecting parts, all assembled on a base plate with a removable hold down bracket, ready for use when equipped with a noncomponent catheter. A package of 24 vacuum container liners, and operating and maintenance manual(s) shall also be included.

The hand pump shall provide suctioning when operated in either direction (push-pull), and shall have a minimum displacement of 12.5 cubic inches (0.205 liters) per stroke.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The collection/vacuum container, with a liner installed, shall have a minimum capacity of 2300 cc and the contents of the liner shall be clearly viewable during use (i.e. installed liner shall also have minimum capacity of 2300 cc). The container liner shall have graduation markings.

The vacuum gauge shall have a range from 0 to a minimum of 30 inches (63.5 cm) Hg.

The vacuum adjustment control shall be hand operated and easily accessed. It shall be capable of controlling the vacuum from 0 to 28 inches (63.5 cm) Hg. The high limit of free air flow shall be 38.5 LPM.

The patient suction tube shall be 1/4 inch (0.635 cm) ID by a minimum of 5 feet (1.52 meters) and shall be sterilizable via ethylene oxide and liquid agent(s).

The vehicle vacuum hose shall be a minimum 16 feet (4.88 meters) long by at least 1/4 inch (0.64 cm) ID.

The complete apparatus, less packaging, package of liners, and manual(s) shall not weigh more than 7 pounds (3.18 kg).

Performance. When manually operated at the rate of two double strokes (2 full cycles) per 2.4 seconds, the apparatus shall provide a minimum of 27.0 inches (68.58 cm) Hg negative pressure with the patient tubing occluded and a minimum of 35 LPM free air flow at the patient end of the non-occluded tubing. When cycle rate is increased, the free air flow shall be limited to a minimum of 38.5 LPM.

When operated from a noncomponent vacuum source (vehicle vacuum or electrical vacuum pump) which has an excess of vacuum and free air flow, the vacuum adjustment control shall provide a range from 0 to a minimum of 27.5 inches (69.85 cm) Hg. The apparatus shall be capable of providing a minimum of 30 LPM free air flow at the patient end of the tube.

Materials. All parts shall be corrosion-resistant or treated for corrosion-resistance (paint is not acceptable as a corrosion-resistant treatment).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Instructions. The apparatus shall be provided with manual(s) describing the theory of operation, operating instructions, assembly and disassembly instructions, a list of replacement parts by manufacturer's catalog number, descriptive sketches and appropriate, statement of warranty, and manufacturer's name, address and phone number.

Workmanship. The suction apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One suction apparatus, as specified, constitutes one unit. Each suction apparatus shall be provided with operating and maintenance instructions for use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. All tests listed under Tests and required by ASTM F 960 shall be performed.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one suction apparatus as specified in this description.

Defects.

Unit not complete with all component parts.

Components not free from damage including (but not restricted to) cracks, checking, breaks, and deformed areas of the collection containers, and tubing.

Unit not clean.

Finish not free from blistering, checking, peeling, scratching, or bare areas.

Any screws not drawn up tight; rivet not properly installed.

Service data not included.

Metal parts not free from corrosion.

Hand pump is not double cycle.

Sampling for tests. Sampling for tests, including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one suction apparatus.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Tests

Measure the length of a stroke of the hand pump. Sufficiently disassemble the pump to measure the stroke area. Calculate the displacement of a single stroke and compare to specified displacement.

Measure the amount of water that the collection container (with a liner in place) is capable of holding and compare with the amount specified.

Test the vacuum gauge for range and accuracy as specified.

Measure the ID and length of the patient suction tube.

Measure the ID and length of the vehicle vacuum hose.

Strenuously operate the hand pump. Measure the free air flow and negative pressure and compare with specifications.

Provide an external vacuum source in great excess to apparatus rating. Set the vacuum limiting control in accordance with instructions. Measure the free air flow and the negative pressure and compare results with limits as specified. Determine if the vacuum limiting control is functioning as specified.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-061-0662

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4372

Location:
CID/SUCTION/APP/H12-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53459	SUCTION APPARATUS, SURGICAL

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for u submitting comments and suggestions for improvements All users of military standardization documents are invited to pro suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpre' too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviat problems Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorizatio to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTA NECESSA IF MAILE IN THE UNITED ST

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4958 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.