
FED. SUP CLASS
6640



PART NUMBER	APPROXIMATE DIMENSIONS						
	CASE inches			LENS inches	COUNTING PLATE millimeters		
	LENGTH	WIDTH	HEIGHT	DIAMETER	LENGTH	WIDTH	THICKNESS
MS36316-1	10	10-3/4	10-1/4	4-1/2	114	114	3.2

MATERIAL: Materials shall be in accordance with the procurement specification.

FINISH: The case shall have a wrinkled-enamel finish.

DESIGN: The counting apparatus shall consist of a lighted case, lens, counting plate and auxiliary tilting base.

The front of the case shall slope at an angle that will provide accurate and comfortable observation of a petri dish placed over the counting plate.

The auxiliary base shall be adjustable to increase the degrees of slope of the front surface of the apparatus if desired.

The support ledge shall have two thumb screws for centering petri dishes over circularly lined plates.

Illumination of the counting plate shall be uniform, bright and free from glare. The lamp shall be 40 watt, easily replaceable.

The lens shall have a magnification of 1.5x and shall be so mounted that it may be positioned from 1-1/4 to 6-1/4 inches from the counting plate or moved aside when not in use.

The counting plate shall be square and have the Wolffhuegel ruling.

NOTES:
1. For Wolffhuegel counting plate replacement see MS36317-1.
2. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 28 JUN 1968 REVISED

NOTICE.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. MU Other Cust BuMed MAA	TITLE COUNTING APPARATUS, BACTERIAL COLONY Quebec, darkfield	MILITARY STANDARD MS36316	
PROCUREMENT SPECIFICATION GG-C-611	SUPERSEDES:	SHEET 1 OF 1	

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:38Z
Check the source to verify that this is the current version before use.