
FED. SUP CLASS
6640



PART NUMBER	TEMPERATURE RANGE °C	[illegible]	ING H2O n)	HEATING CAPACITY W (max)
MS364[illegible]				450
				750
				750
				1000
				1000

CANCELLED
CANCELLED AFTER 23 OCT 1991
NO SUPERSEDING STANDARD

MATER. [illegible] stainless steel in accordance [illegible]-plated copper. All other compo- [illegible] that will function as required for reli-

DESIGN: [illegible]or shall be capable of being mounted on any suitable container [illegible]oratory stand. The temperature control shall accurately control by direct dialing any desired temperature on a continuously variable basis. All circulators have built-in pumps for temperature regulation. For ranges below ambient temperatures accessory cooling coils must by used.

All circulators shall be supplied with a 2 m 3-wire cord and plug, for 115 volts, 50/60Hz power supply.

All electrical components shall conform to the standards of Underwriters Laboratories, Incorporated.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

Ⓐ CANCELLATION

APPROVED 30 JUN 1983 REVISED Ⓐ 23 OCT 1991

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust. MS 99		CIRCULATOR, CONSTANT TEMPERATURE	MS36411
Procurement Specifications NONE	SUPERSEDES		SHEET 1 OF 1

DD FORM 1 NOV 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE CANCELLATION

Project Number 6640-1736-

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

USER ACTIVITIES: ARMY - MD, GL
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:40Z
Check the source to verify that this is the current version before use.