MILITARY SPECIFICATION

COVER, SURGICAL INSTRUMENT AND DRESSING CABINET

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification defines a cover which provides a working area when fitted to the top of a surgical instrument and dressing cabinet.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Box, Shipping, Fiberboard.

MILITARY

MIL-P-116	- Preservation-packaging, Methods Of.
MIL-L-10547	- Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-NO. 66 - Steel; Chemical Composition and Hardenability.
FED-TEST-METHOD-STD-NO. 151 - Metals; Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-810 - Environmental Test Methods.

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

4061 - 3592 - COVER, Lower Cabinet

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING & MATERIALS (ASTM)

ASTM D 1781 - Climbing Drum Peel Test for Adhesives.
ASTM D 1876 - Peel Resistance of Adhesives (T-Peel Test).

(Application for copies should be addressed to the American Society for Testing & Materials, 1916 Race Street, Philadelphia, PA 19103.)

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

No. 56A - Standard for the Use of Inhalation Anesthetics.

(Application for copies should be addressed to the National Fire Protection Association, 470 Atlantic Avenue, Boston, MA 02210.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Materials.

3.2.1 Corrosion-resisting steel. Corrosion-resisting steel shall comply with AISI 300 series of Fed. Std. No. 66 when tested as specified in 4.4.1.

3.3 Design and construction. The cover shall conform to DPSC Drawing No. 4061 - 3592 and as specified herein. The cover shall consist of a corrosion-resisting steel body, to which a honeycomb panel is attached by a suitable epoxy and a corrosion-resisting steel z-frame. Plastic glides shall be supplied to prevent the cover from marring the top of the cabinet. The weight of the cover shall not exceed 14 pounds.

3.4 Electrical conductivity. The completely assembled item shall meet the requirements of NFPA 56A when tested as specified in 4.4.2.

3.5 Bond strength.

3.5.1 Honeycomb panel. When honeycomb panels are tested as specified in 4.4.3, the average peeling torque shall be not less than 7 inch-pounds per inch prior to the humidity test, and not less than 6 inch-pounds per inch after the humidity test.

3.5.2 Top-of-cover to honeycomb-panel. When tested as specified in 4.4.4, the T-peel strength of a sample specimen, bonded under conditions identical to those used in the fabrication of the covers, shall be not less than 7 pounds per inch of width.

3.6 Finish. Exposed corrosion-resisting steel surfaces shall have a commercial standard No. 4 finish.

3.7 Identification marking. Each unit shall be marked permanently and legibly, on the underside of the cover, with the following:

Contract or purchase order number.
Name or registered trademark of manufacturer.
National Stock Number.

3.8 Workmanship. Workmanship shall be first class throughout. Cover shall be free from defects which detract from its appearance or which impair its serviceability. All exposed burrs and sharp edges shall be removed from the cover.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

3.8.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive spatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks. Spot welds shall be smoothly dressed and polished to blend with adjacent surfaces.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the cover when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government and provided also that the test methods and the results obtained comply with those described in this specification. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:21Z
Check the source to verify that this is the currentversion before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling pruposes shall be one cover.

TABLE I. Sampling for examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, for an acceptance level of zero.

TABLE II. Sampling for tests.

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Corrosion-resisting steel	Chemical composition	3.2.1	4.4.1	1/
Cover	Electrical conductivity	3.4	4.4.2	100%
	Bond strength	3.5.2	4.4.4	2/
Honeycomb panel	Bond strength	3.5.1	4.4.3	3/

1/ A sufficient quantity of each lot of each metal used in the manufacture of the cover shall be selected for chemical analysis.
2/ At least two specimens shall be prepared for bond strength testing.
3/ At least six specimens, two or more from each of three honeycomb panels, shall be tested for bond strength.

4.3 Examination. The covers shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:21Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with, but not limited to, the classification of defects listed in table III.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Any component missing, broken, or not as specified.
102	Any weld not as specified.
103	Any bond not as specified.
104	Exposed edge or surface has burrs or sharpness.
105	Any surface dented, bent, or distorted.
Minor	
201	Identification marking not present, complete, legible, permanent, or securely attached.
202	Any rivet not properly installed.
203	Any plastic glide not properly installed.
204	Cover has dirt, grease, or other foreign matter.

4.3.2 Dimensional examination. The cover shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.3.3 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Four ounces from each lot of corrosion-resisting steel shall be tested in accordance with FED-TEST-METHOD-STD-NO. 151.

4.4.2 Electrical conductivity. The completely assembled unit shall be tested using the equipment described in NFPA No. 56A. The test shall be conducted with one electrode attached to the top of the cover, and the other electrode attached alternately to each of the three sections of the z-frame.

4.4.3 Bond strength of honeycomb panels. Six samples, two from each of three honeycomb panels identical to those used in the construction of the covers, shall be used. One sample from each panel shall be tested "as received" as specified in the ASTM D 1781. A separate sample from each panel shall be given the humidity test procedure specified in Method 507, Procedure I, of MIL-STD-810 and then tested as specified in ASTM D 1781. For this test, the combination of the force required to bend the aluminum facing and the force required to cause failure of the adhesive or rupture of the paper core shall be defined as "bond strength".

Source: https://assist.dla.mil -- Downloaded: 20260-21T11:21Z
Check the source to verify that this is the current version before use.

4.4.4 Bond strength of top of cover to honeycomb panel. Two specimens shall be prepared for testing in accordance with ASTM D 1876. Each specimen shall consist of one sheet of anodized aluminum, of the type and thickness used in the honeycomb panel, and one sheet of corrosion-resisting steel, of the type and thickness used in the body of the cover. The adherent sheets shall be bonded in a manner identical to those used in the fabrication of the covers.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning. Each cover shall be cleaned in accordance with MIL-P-116, process C-1.

5.1.1.2 Drying. Each cover shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservation application. Preservatives shall not be used.

5.1.1.4 Unit packaging. Each cover shall be wrapped in kraft paper, cushioned with fiberboard pads or other suitable material and packaged in a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.1.2 Level C. Covers shall be clean, dry and individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.2).

5.2.1 Level A. Four unit packages shall be packed in a shipping container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the box specification.

5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Four unit packages shall be packed in a shipping container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The packaged covers shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked as specified in MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Shipping container. Shipping containers shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The cover is intended for use in field Medical Units, Self-contained, Transportable (MUST). It is suitable for use with cabinets NSN 6530-00-935-9838 and 6530-00-935-9839.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of the Defense Acquisition Regulations. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-110-5547	COVER, SURGICAL INSTRUMENT AND DRESSING CABINET, Corrosion-Resisting Metal.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing activity:
DLA-DM

Project No. 6530-1427

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.
NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-C-37722 COVER, SURGICAL INSTRUMENT AND DRESSING CABINET

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*	TELEPHONE NO.
	DATE

DD FORM 1426 1 OCT 76 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.