MILITARY SPECIFICATION

EVACUATOR, SURGICAL WOUND

This specification is approved for use by Defense Logistics Agency-DM, Department of Defense, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of sterile, disposable surgical wound evacuator, suitable for use in surgery to aspirate fluids and concurrently assist soft tissue separated by surgery or trauma to be drawn together.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, for a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

PPP-B-636 - Box, Shipping Fiberboard.

MILITARY

MIL-S-36586 - Sterilization Strip Test, Bacterial Spore.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)
No. 24390 - EVACUATOR, SURGICAL WOUND.

(Copies of specifications, standards, drawing, and publications, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply:

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (USP).

(Application for copies shall be addressed to the Mack Publishing Company, Easton, Pennsylvania 18042).

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204).

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 Material. Shall be made of high quality plastics, natural rubber and stainless steel. Shall be sterile throughout.

3.1.1 Plastic tubing. The perforated tube shall be made of polyvinyl chloride and be non-pyrogenic and comply with USP Class IV plastics requirements in accordance with 4.4.10 and 4.4.11. The tube shall also contain no more than 5 parts per million (ppm) of lead when tested as specified in 4.4.7 or 4.4.8. The perforated tube shall demonstrate tensile strength and elongation of 8 lbs and 75% respectively, when tested as specified in 4.4.1. The perforated tube shall have a radiopaque line running along its entire length. The perforated tube shall also have a light coating of medical grade silicone oil applied to the exterior surface. The evacuator (non-perforated) tube shall be made of polyvinyl chloride plastic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

3.1.2 Y-Connector. The Y-connector shall be made of plasticized polyvinyl chloride.

3.1.3 Needle. The needle shall be made of austenitic stainless steel and have a mirror finish. The needle shall be lightly coated with medical grade silicone oil.

3.1.4 Evacuator. The evacuator vessel shall be rigid and made of transparent ABS plastic. The diaphragm or gland, shall be made of natural gum rubber and coated with medical grade silicone on the top surface. The plate top, the floating plate and the handle shall be made of white ABS plastic. The valve drain and plug shall be made of polyvinyl chloride plastic colored red. The connector shall be made of clear polyvinyl chloride.

3.1.5 Straps. Two attaching straps shall be made of polyvinyl chloride. Another strap shall be made of polyethylene.

3.2 Style, design and construction. The evacuator shall be a lightweight, portable pump which can be primed and left in this condition when handed to the surgeon by the nurse. The surgeon can then connect the tubing, insert the plug, trip the handle and allow evacuation to proceed. There shall be no need for the nurse to hold the evacuator in the primed position with two hands while the surgeon attaches the tubing and inserts the plug. The contents of the evacuator as it fills shall be easily visible through the clear ABS vessel. The evacuator, tubing, connector and needle shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24390. The attaching straps shall be as follows:

Two - PVC, about 1-1/2 x 12 x .015
One - Polyethylene, about 1 x 60 x .003

3.3 Breaking strength and ultimate elongation. The breaking strength of the perforated tube, including all perforations, shall be a minimum of 8 pounds when tested as stated in 4.4.1. The ultimate elongation for the same portion of perforated tube, including all perforations, shall be 75% when tested as stated in 4.4.1.

3.4 Security of attachments. The attachments shall have the following minimum force separation requirements when tested in accordance with 4.4.2:

Unperforated tube to vessel connector	1.0 lb minimum
Perforated tube to needle	7.5 lbs minimum
Unperforated tube to Y-connector	1.0 lb minimum
Perforated tube to Y-connector	1.0 lb minimum

3.5 Aspiration volume. The evacuator shall aspirate 400 cubic centimeters (cc) of water when tested as specified in 4.4.3.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:21Z
Check the source to verify that this is the current version before use.

3.6 Volume accuracy. The accuracy of the evacuator shall be determined at 200 cc and 400 cc in accordance with 4.4.4. The accuracy shall be 10% or 25 cc whichever is less.

3.7 Initial suction. The initial suction shall be not less than 50 milimeters mercury (mm Hg) when tested in accordance with 4.4.5.

3.8 Leakage of evacuator. The evacuator shall be tested for leakage in accordance with 4.4.6.

3.9 Lead in tubing. The perforated tube shall be tested for lead content in accordance with 4.4.7 or 4.4.8. Lead if present shall not exceed 5 parts per million (ppm).

3.10 Sterility. The evacuator and related accessories shall be sterile when tested with a biological indicator as specified in 4.4.9.

3.10.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, incculated product, or incculated simulated product).
i. Test organisms and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of tests units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.10.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simutaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

3.11 Class IV, USP. The perforated tubing shall be in compliance with the requirements for USP Class IV plastics as indicated in 4.4.10.

3.11.1 Certificate of compliance, USP Class IV. A certificate of compliance with USP Class IV plastics requirements, listing all lot or control numbers, shall state that samples from each lot or control have been tested for USP Class IV as specified in 3.11 and were found to comply with such requirements. The certificate shall reference the methods used in determining compliance with the requirements for USP Class IV plastics.

3.12 Non-pyrogenicity. The perforated tube shall be non-pyrongenic when tested as specified in 4.4.11.

3.12.1 Certificate of freedom from pyrogens. A certificate of freedom from pyrogens, listing all lot or control numbers, shall state that samples from each lot or control have been tested for pyrogens as specified in 3.12 and were found to comply with such requirements. The certificate shall reference the method used in determining freedom from pyrogens.

3.13 Instructions. A detailed instruction sheet shall be provided in order to indicate proper usage.

3.14 Workmanship. Workmanship shall be first class throughout. The evacuator and accessories shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:21Z
Check the source to verify that this is the currentversion before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the plastics the rubber, or the steel, may be furnished in lieu of actual performance of such testing by the contractor provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I. The unit of product for examination purposes shall be one evacuator complete with accessories.

TABLE I. Sampling criteria for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	I	1.0
Minor	I	2.5

4.2.2 Sampling for tests. Sampling for tests shall be in accordance with MIL-STD-105 and table II. The unit of product for tests shall be one wound evacuation with accessories.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Requirement	Procedure	Inspection level	AQL (percent defective)
Perforated tube	Breaking strength	3.3	4.4.1	S-1	1.0
Perforated tube	Ultimate elongation	3.3	4.4.1	S-1	1.0
Evacuator complete with accessories	Security of attachments	3.4	4.4.2	S-2	1.5
Evacuator with tubing & Y-connector	Aspiration volume	3.5	4.4.3	S-2	1.0
Evacuator	Volume accuracy	3.6	4.4.4	S-1	1.0
Evacuator	Initial suction	3.7	4.4.5	S-1	1.0
Evacuator & non-perforated tube	Leakage of evacuator	3.8	4.4.6	S-2	1.0
Perforated tubing	Lead	3.9	4.4.7 or 4.4.8	S-1	1.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling criteria for tests. (cont'd)

Component	Characteristic	Requirement	Procedure	Inspection level	AQL (percent defective)
Evacuator complete with accessories	Sterility	3.10	4.4.9	S-3	1/
Perforated tubing	USP Class IV	3.11	4.4.10	S-2	1.0
Perforated tubing	Non-pyrogenicity	3.12	4.4.11	S-1	1/

1/ Acceptance number shall be zero.

4.3 Examination. The evacuator complete with accessories shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include but shall not be limited to the defects listed in table III.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Perforated tube not smooth and not free from burrs.
102	Perforated tube-needle junction not smooth, i.e., lacking uniformity of diameters at junction.
103	Lack of concentricity of perforated tube and needle at junction.
104	Instructions mission or illegible.
105	Trocar point not sharp.
106	Printing or components flaking off.
107	Presence of foreign matter in or on perforated tube.
Minor	
201	Tubing not free of discoloration.
202	Presence of foreign matter in or on evacuator.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Breaking strength and ultimate elongation. The breaking strength and ultimate elongation of the perforated tube shall be performed on a tensile testing machine having a jaw speed of 20 ± 2 inches/minute.

4.4.2 Security of attachments. Weights as specified in 3.4 shall be applied gradually to test each of the attachments. The load shall be maintained for a minimum of ten seconds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

4.4.3 Aspiration volume. The evacuator, non-perforated tube, Y-connector and perforated tube shall all be connected. A large pan of water shall be present with the water level equal within one-half inch to that of the rubber diaphragm or gland in the fully primed state, i.e., the bottom of the evacuator vessel.

4.4.4 Volume accuracy. The volume shall be determined at the 200 cc and the 400 cc levels by filling the evacuator to these levels and then emptying the evacuator contents into a graduate.

4.4.5 Initial suction. The evacuator shall be fully primed, connected to a manometer, plugged and the handle released. The negative pressure shall be read from the manometer.

4.4.6 Leakage of evacuator. The non-perforated tube shall be attached to the evacuator. The drain port shall be plugged. The end of the tube shall be connected to a vacuum pump and air aspirated to a vacuum of 160 mm Hg. The evacuator at the negative pressure of 160 mm Hg shall then be immersed in water for 10 seconds and observed for any sign of leakage.

4.4.7 Test for lead. Ten (10) Gms of rubber or plastic shall be taken, ignited to char, and then further ignited in a muffle furnace at 700° C. to a white ash. The ash shall be boiled in 15 milliliters (ml) of 10 percent nitric acid, filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead as given in the Pharmacopeia of the United States, shall be applied using a control of 5 ppm or 12 ppm of lead solution as given in the USP test. If the sample solution is darker than the limit, small aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in ppm is determined.

4.4.8 Alternate test for lead. Accurately weigh a 5-gram sample into a 1-oz. Vycor crucible. Reduce to ash in a muffle furnace at 650° C for eight hours. Let cool to room temperature and add 15 ml. of 1:1 hydrochloric acid. Bring to a boil on a hot plate, allow to boil for 2 to 3 minutes, remove, let cool, and filter through Whatman #2 filter paper into a 50-ml volumetric flask. Rinse the residue on the filter with deionized water and dilute to 50 ml. Prepare standard lead solutions of 1, 2 and 5 ppm lead. Determine the absorbances of these solutions at the 283 millimicron wave length using a Hollow Cathode Pb lamp and a 10x scale expansion on a spectrophtometer, such as the Perkin-Elmer Model 303 Atomic Absorption Spectrophotometer. Determine the absorbance of the sample solutions using the same conditions as with the standards.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

Calculation:

$$\frac{A \times F \times \text{dilution}}{W} = \text{ppm lead wherein,}$$

A = Absorbance
F = Factor = $\frac{\text{(Concentration standard in ppm Pb)*}}{\text{absorbance}}$
W = Sample weight in grams

* The three standards are averaged to give ppm/unit A.

4.4.9 Sterility test. The evacuator complete with accessories shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.9.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in table II, Sampling Criteria for Tests.

4.4.9.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. substilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.9.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.9. Sterility test and 4.4.9.1 Inoculated product or simulated product. The method of sterilization will control the one organism in the spore strip to be tested for as shown in 4.4.9.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

4.4.9.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.9.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.10 USP Class IV plastics. The perforated tubing shall be tested according to the methods described in the USP for Class IV plastics except that implant time shall be 72 to 144 hours and the muscle tissue surrounding the implants shall be examined both grossly and microscopically by a certified pathologist or a pathologist certified by the American College of Veterinary Pathologists.

4.4.11 Freedom from pyrogens. Samples of the perforated tube shall be tested for freedom from pyrogens in accordance with the USP for Transfusion and Infusion assemblies.

4.5 Packaging inspection. The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. All components comprising the evacuator shall be thoroughly clean and dry. The tubing shall be neatly coiled and secured with paper band. The needle shall have the end protected with a PVC sleeve. The coiled tubing, needle, connector and straps shall be placed in a 3 mil polyetheylene bag. Closure of the bag shall be adequate to prevent spilling of contents under normal handling. The bagged components along with the evacuator shall be packaged in a sealed, peel-open type bag capable of maintaining sterility of contents unless damaged or opened. The sealed bag shall be further packaged in a paperboard box of an appropriate size and design. Closure of the box shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

5.1.2 Level C. Each evacuator shall be clean, dry and packaged in a manner that will maintain sterility and afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Ten unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Ten unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged evacuators shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking.

5.3.1 Unit package.

5.3.1.1 Peel-open bag. The bag shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. The lot (control) number and any other markings usually furnished by the supplier as his commercial practice shall be shown. The legend: STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement shall also be shown.

5.3.1.2 Unit box. The box shall be marked as specified in MIL-STD-129. The lot control number shall be shown. The method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry heat sterilization) shall be shown following the level of packaging and date of packaging: Example: "A 1/77-G". Final sterilization shall be on or after indicated packaging date.

5.3.2 Exterior container. Exterior containers shall be marked as specified in MIL-STD-129. Lot (control) number and the method of sterilization as specified in 5.3.1.2 shall be shown.

6. NOTES

6.1 Ordering data.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. National Stock Number.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:21Z
Check the source to verify that this is the current version before use.

6.2 National stock number coverage. This specification covers the following item(s) appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-149-0671	EVACUATOR, SURGICAL WOUND, Disposable, 400 cc

Preparing activity:
Defense Logistics Agency-DM

Project No.. 6515-2088

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-E-37718 (-) EVACUATOR, SURGICAL WOUND

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.