# MILITARY SPECIFICATION

## LIGHT, SURGICAL, CEILING

This specification is approved for use by all Departments and Agencies of the Department of Defense

1 SCOPE

1.1 Scope. This specification covers a surgical ceiling light, 24 volts, DC, suitable for use in surgical facilities of military field hospitals, and a case to protect the reflector assembly

2 APPLICABLE DOCUMENTS

2.1 Issues of documents The following documents of the issue in effect on date of invitation for bid or request for proposal, form a part of this specification to the extent specified herein

SPECIFICATIONS

FEDERAL

| | |
|---|---|
| TT-E-490 | - Enamel, Gloss, Synthetic (for Metal and Wood Hospital Furniture) |
| TT-P-1757 | - Primer Coating, Zinc Chromate Low Moisture Sensitivity |
| PPP-B-601 | - Boxes, Wood, Cleated-Plywood |
| PPP-B-621 | - Boxes, Wood, Nailed and Lock-Corner |
| PPP-B-636 | - Boxes, Shipping, Fiberboard |

MILITARY

| | |
|---|---|
| MIL-F-495 | - Finish, Chemical, Black for Copper Alloys |
| MIL-W-6858 | - Welding, Resistance, Aluminum, Non-hardening Steels or Alloys, Nickel Alloys, Heat-resisting Alloys, and Titanium Alloys, Spot and Seam |
| MIL-M-7298 | - Manual, Technical Commercial Equipment |
| MIL-L-10547 | - Liners, Case and Sheet Overwrap Water-Vaporproof or Water-proof, Flexible. |
| MIL-P-15328 | - Primer (Wash), Pretreatment, Blue (Formula No 117-E for Metals) |

Beneficial comments (recommendations, additions, deletions and any pertinent data which may be of use in improving this document should be addressed to Defense Personnel Support Center, Directorate of Medical Materiel DPSC-ATT 2800 South 20th Street, Philadelphia, PA 19101 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426 appearing at the end of this document or by letter

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-TEST METHOD-STD-NO. 151 - Metals, Test Methods
FED-STD-595 - Colors
FED-STD-751 - Stitches, Seams, and Stitchings

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129 - Marking for Shipment and Storage

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

25039 - Light, Surgical, Ceiling, 24 V., D.C.
3061-4500 - Light Transit Case

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING & MATERIALS (ASTM)

B-136 - Measurement of Stain Resistance of Anodic Coatings of Aluminum

(Application for copies should be addressed to the American Society for Testing & Materials, 1916 Race Street, Philadelphia, PA 19103.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Material. Except as hereinafter specified, materials shall conform to the applicable portions of the specifications, standards, and drawings specified herein. Materials shall be subjected to the tests and meet the requirements specified in these documents. The metals shall conform to the chemical and physical requirements of the applicable specifications. Materials, not specifically designated herein, but used in the fabrication of the lights, shall be suitable for the purpose intended.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

3 3 Design and construction

3 3.1 Light The light shall be constructed in accordance with DPSC Drawing No 25039 When mounted against a flat, level ceiling, the bushing at the back of the reflector shall be not less than 1/2 inch from the ceiling nor more than 4-1/2 inches from the ceiling, with the reflector in the full up position and pointing straight down The reflector shall have a minimum vertical travel range of 12 inches down from the ceiling and the light assembly shall rotate a full 360 degrees around the mounting base The inner (reflecting) surface of the reflector shall be given a specular finish protected by a clear anodic coating at least 0 0001 inch thick (see 6.4) to protect against clouding, staining, oxidizing, and peeling The reflectivity shall be at least 83 percent The reflecting surface shall show no evidence of stain or discoloration when tested as specified in 4 5 6 A heat-dissipating glass cylinder, capable of absorbing more than 90 percent of the infrared radiation, shall be provided An extra lamp shall be supplied

3 3 1 1 Cable length The power cable (part no 42) shall be of sufficient length to provide adequate strain relief in all adjustments of the light, and there shall be 24, + 1-1/2 -1/2 inches of cable between the strain-relief bushing (part no 89) and the plug (part no 65)

3 3 1 2 Polarity

3 3 1 2 1 Plug The plug (part no 65) shall be wired as follows

| Conductor color | Function | Pin number |
|---|---|---|
| Gray | 0 volts, D C | A-1 |
| Green | Local ground | A-2 |
| Black | +28 volts D C | A-3 |

3 3 1 2 2 Lamp holder The lampholder (part no 37), located in the reflector, shall be wired as follows

| Conductor | Function | Terminal |
|---|---|---|
| Black | +28 volts, D C | Brass (center contact) |
| Gray | 0 volts D C | White (screw shell) |
| Green | Local ground | Metal frame |

3 3 2 Case The protective case for the reflector assembly shall be constructed in accordance with DPSC Drawing No 3061-4523 Tolerance on linear dimensions shall be ± 1/4 inch unless otherwise specified

3 3 2 1 Stitching All stitching except bartacking shall conform to stitch type 301 of Fed Std 751 with a minimum of 8 stitches per inch Thread breaks in stitching shall be backstitched not less than inch at each break Stitching shall be backstitched not less than 1 inch at the ends except where ends are turned under in a hem or held down by other stitching or on box stitching, where it shall be not less than 3/8 inch Thread tension shall be maintained so that there will be no loose stitching and so that the lock will be [illegible] All thread ends shall be trimmed and removed As an alternate, back stitching pattern as performed by box stitching machines is acceptable Bartack stitching shall be 5 8 to 3/4 inch in length and shall contain 38 to 42 stitches Bartacking shall be free of thread breaks or loose stitching.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

3.3.2.2 Handles. Handles shall be fabricated of webbing and securely stitched around all edges in contact with case Bartack stitching shall be used to secure the handles to the support strap inside the case Raw ends of webbing shall be turned under and double stitched

3.3.2.3 Support strap Support strap shall be secured by stitching around the edges.

3.3.2.4 Cover All cover seams shall be sewn with two rows of stitching, ½ inch apart

3.4 Performance.

3.4.1 Illumination The light shall produce a minimum of 1500 foot candles at a point on the axis of the reflector and 42 inches from the front edge of the reflector when tested as specified in 4.5 3

3.4.2 Color temperature The color temperature of the light shall be $4000^{0}$ ± $500^{0}$ Kelvin when tested as specified in 4 5 4.

3 4 3 Temperature rise The temperature of a blackened disc shall not rise more than $10^{0}$C above ambient when tested as specified in 4 5 5

3.5 Finish Unless otherwise specified, all exposed metal parts of the light shall be treated with a wash primer coat in accordance with MIL-P-15328 and then treated with a primer coat in accordance with TT-P-1757 If required, a filler coat of paint shall be applied to the cast aluminum arm to create a more uniform and even finish to the casting The primer coat (or the filler coat, if used) shall be followed by two coats of Class B, gloss enamel in accordance with TT-E-491 Color shall be gray No 16473 of FED-STD-595 The following components shall not be painted

a Reflecting surface
b Electrical components
c. Sterilizable handle assembly.
d Inner components of arms
e Hardware made of corrosion-resisting material or with corrosion-resisting finish

3 6 Dielectric withstand The light shall show no evidence of dielectric breakdown when tested as specified in 4 5.2

3.7 Identification marking

3.7.1 Light Each light shall be marked, permanently and legibly with the name or registered trademark of the manufacturer, contract or purchase order number, National Stock Number, and electrical characteristics (volts, amperes D C , and watts) Marking shall be by means of steel stamping or a metal nameplate, permanently attached to the light arm.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

3.7.2 Case. The exterior of the case shall be marked, permanently and legibly, in black indelible ink, in capitals 1 inch high, with the following

6530-00-027-5018
LIGHT, SURGICAL CEILING,24 VOLT, D C
MANUFACTURER'S NAME
CONTRACT OR PURCHASE ORDER NUMBER

3.8 Service data. The contractor shall furnish with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical to those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.9 Workmanship. Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or impair its serviceability. All burrs and sharp edges shall be removed from the item.

3.9.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks. Welding of aluminum alloys shall be in accordance with MIL-W-6858.

4 QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

4.1.2 Inspection Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection The inspection requirements specified herein are classified as follows

a. First article inspection
b Quality conformance inspection

4.1.2.1.1 First article inspection First article inspection shall be performed on the light when a first article sample is required (see 3.1) This inspection shall include examination of 4.4 and tests of 4.5

4.1.2.1.2 Quality conformance inspection Quality conformance inspection shall include the examination of 4.4 and tests of 4.5

4.1.3 Certificates of quality Certificates of quality supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government and provided also that the test methods and the results obtained comply with those described in this specification The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements

4.2 Inspection lot.

4.2.1 Raw materials An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of product within the designated lot can be assured. Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable

4.2.2 End product An inspection lot for inspection of end product shall consist of all lights of the same type produced under essentially the same conditions and offered for inspection at one time



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

4 3 Sampling

4 3 1 For examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I The unit of product for sampling purposes shall be one light, including the case

TABLE I Sampling for examination

| | Inspection level | AQL (percent defective) | |
|---|---|---|---|
| | | Light | Case |
| For visual examination | | | |
| Major defects | II | 1 0 | 4 0 |
| Total defects (Major and Minor combined) | II | 4 0 | 10 0 |
| For dimensional examination | S-1 | 2 5 | 2 5 |

4 3 2 For tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, for an acceptance level of zero

TABLE II Sampling for tests

| Component (Unit of product) | Characteristic | Requirement | Test procedure | Inspection level |
|---|---|---|---|---|
| Light | Illumination | 3 4 1 | 4 5 3 | 1/ |
| | Color temperature | 3 4 2 | 4 5 4 | 1/ |
| | Temperature rise | 3 4 3 | 4 5 5 | 1/ |
| | Dielectric withstand | 3 6 | 4 5 9 | 100% |
| Reflector | Stain resistance | 3 3 1 | 4 5 6 | 1/ |

1/ One unit from each lot

4 3 3 Raw materials A sufficient quantity of each lot of each metal used in the manufacture of the light shall be selected for chemical analysis

4 4 Examination The light shall be examined to determine compliance with all requirements contained in this specification.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

4 4.1 Classification of defects Examination shall be conducted in accordance with the classification of defects listed in table III Examination shall not be restricted to the classified possible defects listed below

TABLE III Classification of defects

| Categories | Defects |
|---|---|
| | LIGHT |
| Major A | |
| 101 | Any component part missing, broken, or otherwise unusable |
| 102 | Exposed edge has bur or sharpness. |
| 103 | Components not properly aligned and fitted |
| 104 | Any weld cracked, broken, or contains pinholes |
| 105 | Sterilizable handle not easily removed or replaced |
| 106 | Light does not adjust freely throughout specified range |
| 107 | Wiring raceways have burs or sharpness |
| 108 | Strain relief inadequate in any position of adjustment. |
| | CASE |
| Major B | |
| 110 | Any component not as specified |
| 111 | Material has cut, tear, hole or rupture |
| 112 | Stitching not backstitched at least 1/2 inch at each end and each break (unless otherwise specified) |
| 113 | Less than 6 stitches per inch |
| 114 | Foam not cemented securely in place |
| 115 | Slide fastener does not operate freely |
| | LIGHT |
| Minor A | |
| 201 | Finish chipped, cracked, scratched, or rough |
| 202 | Identification marking not complete correct, permanent, and legible |
| 203 | Nameplate not securely attached |
| 204 | Unit not clean has dirt, grease, or oil |
| 205 | Service data not as specified. |
| | CASE |
| Minor B | |
| 211 | Less than 8 stitches per inch (but at least 6) |
| 212 | Stitching backstitched less than 1 inch (but at least 1/2 inch) at any end or break (unless otherwise specified) |
| 213 | Identification marking not complete correct permanent and legible. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

4.4.2 Dimensional examination. The light shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect

4.4.3 Packaging inspection The inspection of the packaging packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129

4.5 Tests Tests shall be conducted to determine compliance with specification requirements

4.5.1 Chemical composition Four ounces from each lot of metal shall be tested in accordance with FED-STD-151.

4.5.2 Dielectric withstand A slowly rising potential, from 0 to 900 volts, 60 Hz, shall be applied between live and dead metal parts This potential shall be maintained at maximum for one minute

4.5.3 Illumination The light shall be completely assembled and attached to the 24 volts, D.C. power supply (see 6.5) With no ambient lighting, the illumination shall be measured at a point on the axis of the reflector, 42 inches in front of the leading edge of the reflector, using a suitable measuring device accurate to within ± 5 percent

4.5.4 Color temperature Color temperature shall be measured at an ambient temperature of 24°C ± 4°C using a suitable measuring device accurate to ± 5 percent The color temperature derived from the chromaticity diagram and certification of the heat cylinder by the manufacturer, and the certification of the lamp by the manufacturer, shall be acceptable in lieu of the actual measurement

4.5.5 Temperature rise A black disc, not less than 4 inches in diameter, shall be placed at a spot on the axis of the reflector, 42 inches in front of the leading edge of the reflector, at an ambient temperature of 24°C ± 4°C A suitable temperature measuring device, accurate to within ± 1°C, shall be attached to the center of the disc. The temperature of the disc shall be recorded before the light is turned on The light shall be turned on and shall remain on for three continuous hours Temperature readings shall be taken every half-hour

4.5.6 Stain resistance The reflecting surface shall be tested for stain resistance as specified in ASTM B-136



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2)

5.1.1 Unit of issue. One light assembly with case or one case as specified, constitutes one unit of issue

5.1.1.1 Level A

5.1.1.1.1 Unit package.

5.1.1.1.1.1 Light assembly The light head, adjusting arms and mounting plate shall be disassembled. The light head and one extra lamp shall be packaged within the protective case. The protective case with light head enclosed along with the disassembled adjusting arms and mounting plate, shall be wrapped in kraft paper or other suitable material and packaged in one or more boxes of appropriate size constructed in accordance with PPP-B-636 type CF, class domestic Boxes shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of or damage to the contents Closure shall be adequate to prevent accidental opening.

5.1.1.1.1.2 Case Each case shall be packaged in a box of appropriate size and design and constructed in accordance with PPP-B-636, class weather resistant, grade V3c Closure, waterproofing and strapping shall be as specified in the appendix of PPP-B-636.

5.1.2 Level C The light assembly or case shall be packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from supply source to the first receiving activity

5.2 Packing Packing shall be level A, B or C as specified (see 6.2)

5.2.1 Level A. Each light assembly, packaged as specified in 5.1.1.1.1 shall be packed in an exterior container constructed in accordance with PPP-B-601, overseas type, or PPP-B-621, class 2 Closure and strapping shall be in accordance with the appendix of the applicable box specification Cases packaged as specified in 5.1.1.1.2 need no further overpacking

5.2.1.1 Case liner Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547 Closure and sealing shall conform to applicable paragraphs of appendix thereto



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

5 2 2 Level C. The packaged lights or cases shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity These packs shall conform to the applicable carrier rules and regulations

5 3 Marking Interior packages specified in 5 1 1 1 1 1 shall be marked to identify contents and quantity contained Exterior container shall be marked as specified in MIL-STD-129 Marking of the unit package specified in 5 1 1 1 1 2 shall be marked as required for the exterior container specified in MIL-STD-129

6 NOTES

6 1 Intended use The light is intended for use in military field hospitals

6 2 Ordering data Purchasers should select the preferred options permitted herein and include the following information in procurement documents

a Title, number and date of this specification
b Selection of applicable levels of packaging and packing (see 5 1 and 5 2)
c When a first article is required (see 3 1, 4 1 2 1 and 6 3)

6 3 First article When a first article is required, it shall be tested and approved under the appropriate provisions of the Defense Acquisition Regulation, Armed Services Procurement Regulation The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article

6 4 An aluminum finishing process that provides the specified reflectivity, thickness and stain resistance is Alzak, which is available from the Aluminum Co of America

6 5 The power supply which is designed for use with the surgical ceiling light is listed in the Federal Supply Catalog under NSN 6530-00-131-6974

6 6 This specification covers the following items appearing in the Federal Supply Catalog

| National Stock Number | Item Identification |
|---|---|
| 6530-00-027-5018 | LIGHT, SURGICAL, CEILING, 24 volt, D C |
| 6530-00-112-9320 | CASE STORAGE, SURGICAL CEILING LIGHT |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

Custodians
Army - MD
Navy - MS
Air Force - 03

Review activity
DoD - MB

Preparing activity
DLA-DM

Project No 6530-1431



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS This form is provided to solicit beneficial comments which may improve this document and enhance its use DoD contractors government activities manufacturers, vendors or other prospective users of the document are invited to submit comments to the government Fold on lines on reverse side, staple in corner and send to preparing activity Attach any pertinent data which may be of use in improving this document If there are additional papers attach to form and place both in an envelope addressed to preparing activity A response will be provided to the submitter, when name and address is provided within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed

NOTE This form shall not be used to submit requests for waivers deviations or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

DOCUMENT IDENTIFIER *(Number,* AND TITLE

MIL-L_37745 LIGHT, SURGICAL, CEILING

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1 ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID RESTRICTIVE LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDED WORDING CHANGE

C REASON FOR RECOMMENDED CHANGES

2 REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO

DATE

DD FORM 1 OCT 76 1426

EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID




DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:25Z
Check the source to verify that this is the current version before use.