MILITARY SPECIFICATION

SINK, CLINICAL LABORATORY, LIGHTWEIGHT

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of double-basin laboratory sink having a hinged cover which also serves as a glassware drying rack (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.

MILITARY

MIL-P-116	- Preservation-packaging, Methods of.
MIL-M-7298	- Manual, Technical, Commercial Equipment.
MIL-A-8625	- Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-L-10547	- Liners, Case and Sheet Overwrap, Water-Vaporproof, Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-TEST METHOD STD-151 - Metals, Test Methods.
FED-STD-595 - Colors.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-401 - Sandwich Constructions and Core Materials General Test Methods.
MIL-STD-810 - Environmental Test Methods.

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

5061-3328 - Sink Cabinet Assembly.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Design and construction. The design and construction shall be in accordance with DPSC drawing 5061-3328 and bill of material thereto. The laboratory sink shall have a corrosion-resisting steel, double basin sink built into a free-standing cabinet. Cabinet panels shall be paper honeycomb with aluminum facings. Overall dimensions shall be nominally 18 inches deep, 36 inches wide and 37 inches high; however, exact dimensions shall be in accordance with the specified drawing. The drain of each basin shall be equipped with a combination corrosion resistant steel strainer-stopper. Each end of the cabinet shall contain a recessed handle for transportability. The cabinet shall contain an enclosed storage space under one basin, and shall be open under the other basin. Both basins shall drain into a common waste line. The waste line shall have a fitting and hose for attachment to Sink Service Unit (see 6.1). Faucets and hoses of the sink service unit shall also provide water to the laboratory sink. The hinged cover of the laboratory sink shall contain pegs on the underside for glassware drying. The weight of the assembled unit, including sink cabinet and cover, shall not exceed 110 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

3.2.1 Plumbing. All waste drain lines and fittings shall be fabricated from corrosion-resistant materials. The waste drain line shall be capable of being drained to prevent water from remaining in any portion of the system while in transit. When tested in accordance with 4.5.3, there shall be no evidence of leakage.

3.3 Bond strength. When the honeycomb panels are tested as specified in 4.5.2 the average peeling torque per inch of specimen width shall not be less than 9 inch-pounds prior to the humidity test, and not less than 8 inch-pounds after the humidity test. These peeling torque values shall include the torque required to bend the panel facing.

3.4 Finish. Corrosion-resisting steel shall have a Commercial Standard Number 4 finish. Where chromium plating is specified, chromium shall be plated over nickel. Nickel plating shall be a minimum 0.0004 inch in thickness and chromium plating shall be 0.00002 to 0.00005 inch in thickness. When anodizing is specified, anodizing shall conform to Type II, Class 1 of MIL-A-8625; anodized parts shall be unpainted. Prior to painting, carbon steel parts shall be phosphatized, aluminum, zinc and cast-iron parts shall be phosphatized or wash primed. A primer coat shall be applied to a dry film thickness of 0.0004 to 0.0007 inch. A color coat shall be applied to a wet thickness of 0.004 to 0.006 inch. A final, clear texture coat, shall then be applied to achieve a standard texture, and the unit allowed to air dry. Color shall be 24516 of FED-STD-595. Paints shall be applied in accordance with the manufacturer's directions (see 6.5).

3.5 Identification marking. The following information shall be permanently and legibly marked on the nameplate:

National Stock Number.
Nomenclature.
Contract or purchase order number.
Name or registered trademark of the manufacturer.

3.6 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item preceding the date

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.7 Workmanship. Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or impair its serviceability. All burrs and sharp edges shall be removed from the item.

3.7.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Welds shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. Welds shall have no visible voids, pin holes, or cracks, to the unaided eye. Welds, except spot welding, shall be continuous.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the sink service unit when a first article sample is required (see 3.1). This inspection shall include examination of 4.4 and tests of 4.5.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.4 and tests of 4.5.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection lot.

4.2.1 Raw materials. An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of product within the designated lot can be assured. Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable.

4.2.2 End product. An inspection lot for inspection of end product shall consist of all sinks of the same type produced under essentially the same conditions and offered for inspection at one time.

4.3 Sampling.

4.3.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one sink.

TABLE I. Sampling for examination.

	Inspection level	AQL (defect per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional defects	S-3	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

4.3.2 For tests. Sampling for tests shall be conducted in accordance with table II. The unit of product for sampling (test 4.5.3.) purposes shall be one sink.

TABLE II. Sampling for tests.

Component (Unit of product)	Characteristic	Requirement	Test procedure	Acceptance criteria
Panel	Bond strength	3.3	4.5.2	(see 4.5.2)
Sink	Plumbing leakage	3.2.1	4.5.3	Inspection level 100 percent (reject all defective units)

4.3.3 Raw materials. A sufficient number quantity of each lot of each metal used in the manufacture of the sinks shall be selected for chemical analysis.

4.4 Examination. The sinks shall be examined to determine compliance with all requirements contained in this specification.

4.4.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Not style and design specified.
102	Not free of cracks, holes, tears.
103	Any component missing, broken or otherwise unuseable.
104	Welding not as specified.
105	Adhesive not properly applied.
106	Drain line joints not properly heat fused.
Minor	
201	Identification marking not present, complete, legible, correct.
202	Finish not free of scratches, mars, chips.
203	Not clean, free of grease, oil, dirt, other foreign matter.
204	Door or cover does not operate properly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects. (Continued)

Categories	
Minor	
205	Latches not as specified.
206	Color not as specified.
207	Service data not complete, missing.
208	Fasteners or rivets not tight, properly applied.
209	Not free of sharp edges.

Note: Periodic in-process inspections shall be made to insure that honeycomb panels have been properly bonded to mating channels and structural extrusions, in accordance with the requirements herein.

4.4.2 Dimensional examination. The sinks shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4.3 Packaging inspection. The inspection of the unit packaging, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and marking requirements of MIL-STD-129.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Chemical composition. Four ounces from each lot of metal shall be tested in accordance with FED-TEST METHOD-STD 151.

4.5.2 Bond strength. Six specimens which are representative of the honeycomb construction being used shall be selected and tested as specified in the Sandwich Peel Test of MIL-STD-401. Two specimens shall be taken from each of three bonded panels for each lot of adhesive. One half of the samples shall be tested in the "as received" condition and the other half shall be tested after subjecting to the humidity test in method 507 of MIL-STD-810, procedure I. In each testing sequence, the test results from the three specimens shall be averaged to detemine conformance to 3.3. The peeling torque values obtained in the test and specified herein shall include the torque required to bend the aluminum facing of the honeycomb panel. The pulling force shall be applied in the L or ribbon direction of the paper core.

4.5.3 Plumbing leakage. The drain line of assembled sink shall be capped or plugged at the outlets. The strainer-stopper shall be placed in the "stop" (or plugged) position in each basin. Each basin shall be filled with water to within one inch of the top. The water shall remain for 5 minutes and any leakage shall be noted. At the end of the 5 minute period

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

the strainer stoppers shall be removed. The drain line shall then be uncapped and the water permitted to drain, until the basins are approximately one half full. The drain line outlets shall be recapped for a 10 minute period. Any evidence of leakage shall constitute failure.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning. Each sink shall be thoroughly clean.

5.1.1.2 Preservative application. Preservatives shall not be used.

5.1.1.3 Service data. Instruction books and parts lists shall be packaged in a clear sealed polyethylene bag and secured to the sink in a protected location.

5.1.1.4 Cover and door. The sink cover and cabinet shall be secured in a closed position by use of tape or other suitable means.

5.1.2 Level C. Each sink shall be clean and packaged in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A or C, as specified (see 6.2).

5.2.1 Level A. Each sink, prepared as specified in 5.1.1, shall be packed in a container designed for a type 3 load and conforming to PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. The sink shall be cushioned within the box by use of fiberboard pads or other suitable material in such a manner so as to prevent movement or damage to the unit during shipment and handling.

5.2.1.1 Case liner. Wood boxes shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the appendix of MIL-L-10547. Paperboard boxes shall be waterproofed as specified in the appendix of the box specification.

5.2.1.2 Closure. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B. Each sink shall be packed in a shipping container designed for a type 3 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. Each packaged sink shall be packed in a shipping container in a manner that will afford adequate protection against damage during shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking. Marking shall be as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The laboratory sink is intended for use in the Medical Unit Self-contained Transportable (MUST) to provide sink facilities and a glassware drying rack. Two laboratory sinks shall mate with one SINK, SERVICE UNIT, CLINICAL LABORATORY, 6545-00-019-9330. The sink service unit shall provide water and drain connectors for the two laboratory sinks. The sinks may be stored and transported in an ambient temperature range of -65°F to 160°F.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information on the procurement documents:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required it shall be tested and approved under the appropriate provisions of the Defense Acquisition Procurement Regulations. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6545-00-027-5009	SINK CLINICAL LABORATORY, LIGHTWEIGHT

6.5 The phosphate treatment shall be Parker Rust Proof Co. Bonderite 100, or equal; wash primer shall be John L. Armitage and Co. P-147, or equal. Primer coat shall be Armitage and Co. P-289, or equal. The color coat shall be Armitage and Co. K-497, or equal. The clear texture coat shall be Armitage and Co. M-516, or equal.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing activity:
DLA - DM

Project No. 6545-0032

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-S-37721 SINK, CLINICAL LABORATORY, LIGHTWEIGHT

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:21Z
Check the source to verify that this is the current version before use.