A-A-30046B
November 27, 1991
SUPERSEDING
A-A-30046A
October 13, 1982

COMMERCIAL ITEM DESCRIPTION

BAG, TUBE FEEDING

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a disposable enteral tube feeding bag for use with patients who require enteral nutrition support.

Salient characteristics. The feeding bag shall be a nonsterile, disposable, individually packaged, clear medical grade plastic bag with a pre-attached administration set suitable for enteral nutrition gravity delivery. The bag is not for use with electronic enteral delivery systems.

The bag shall have a capacity of 1000 ml with clear graduations in 100 ml increments to the 1000 ml level. Accuracy of the graduations and capacity shall be to within + or - 5% when tested by use of a laboratory graduate capable of accurate measurement of fluids in 1 ml increments. The bag shall have a leakproof opening at or near the top for ease in adding fluids. The top of the bag shall also be provided with a device with which to securely attach a filled bag to an IV pole. The bag shall be designed to drain completely, leaving no residual fluids.

The bag and administration set shall not leak when filled to capacity with water.

The administration set shall be not less than 65 inches (165.10 cm) long as measured from the distal connector to where it connects to the bag. The set shall consist of clear plastic tubing with 3/16 inch (0.475 cm) (+ or - 1/32 inch (0.079 cm)) ID, an integral clear plastic drip chamber for flow rate observation, a roller clamp capable of regulating and complete shut-off of flow (the flowrate drift shall not be more than 4 drops per minute off the set rate) and terminate with a permanently attached male

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

step-wise tapered adapter that is capable of leakproof connection to funnel end feeding tubes. The adapter/connector shall be covered with a cap to protect cleanliness of the set until the time of use.

Workmanship. The feeding bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). A single individually packaged feeding bag usually provided by the manufacturer as commercial practice, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Leak test. Fill the bag to graduated capacity of 1000 ml with the roller clamp in the shut-off position. Invert the bag and hang, using the hanging device, so the bag and administration set is fully extended. Open the clamp to allow water to enter the tubing and drip chamber. Close the clamp when the tubing is filled. Add water to the bag as necessary to bring it back to full capacity of 1000 ml. The bag shall remain inverted for 4 hours and then observed for leaking. No leak at any point is allowed.

Roller clamp test. Perform after the leak test. With the bag still hanging and inverted, slowly open the roller clamp to permit drops of fluid to enter the drip chamber. Without further movement of the clamp, count the number of drops entering the drip chamber for 15 seconds. Adjust the clamp to get 15 drops per 15 seconds. Repeat the 15 second count for at least 2 cycles without movement of the clamp. There should be little change in the number of drops entering the drip chamber in these 15 second cycles. Repeat the above by adjusting the clamp to get 20 drops in 15 seconds and repeat as above noting the number of drops in the 15 second intervals. The flowrate drift of the clamp shall not be more than 4 drops per minute off the set rate.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and inspection level of II. Sampling for tests shall be in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit or product for test purposes shall be one feeding bag.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to the first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereafter. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be the degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be that usually provided by the supplier as commercial practice, and as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6515-00-481-2049	BAG, TUBE FEEDING, Plastic, Disposable, Individual Package, 1000 ml capacity.

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes, may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select from the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

Project No. 6515-5088

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.