A-A-53230A
8 June 1989
SUPERSEDING
A-A-0053230(DM)
28 September 1987

COMMERCIAL ITEM DESCRIPTION

TUBE, TRACHEAL, Pediatric (Murphy)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers four sizes of sterile, disposable, uncuffed pediatric Murphy-type tracheal tubes.

Salient characteristics:

Style and design. The style and design of the pediatric tracheal tubes shall be in accordance with American National Standards Institute (ANSI) Standard Z79.14. The size and the respective inside diameter are delineated below:

Size	Inside diameter (millimeters)
2.5	2.50 ± 0.15
3.5	3.50 ± 0.15
4.5	4.50 ± 0.15
5.5	5.50 ± 0.15

The tube shall be supplied with a 15 mm tracheal tube adapter. Style and design of the adapter shall be in accordance with ANSI Standard Z79.2.

Materials. Tracheal tubes shall be fabricated from polyvinyl chloride (PVC). This material shall comply with requirements listed in Table 1. Samples shall be tested to determine compliance with material requirements.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Table I
Material Requirements

Requirement		Test Method
Elastic Modulus:	10,000 psi, Maximum	ASTM D638
Water Absorption:	1% Maximum	ASTM D570*
Shore Hardness:	A50-A100	ASTM D2240
Toxicity:	Non-Toxic	ASTM F895

*Perform the 24- hour water absorption test.

Radiopacity. The tracheal tube shall have a single radiopaque stripe running the entire length of the tube. The tube shall be radiopaque when tested in accordance with Method C of ASTM Designation F640.

Sterility. The tube shall be sterile in accordance with the USP or AAMI requirements. Sterility assurance level shall be 1×10^{-6}.

Workmanship. The tube shall be free from blooming, blisters, fish-eyes, frosting, pits, pimples, shorts, warpage, wrinkles, and other defects that detract from their appearance of impair their serviceability (See Note 2).

Unit. Package (PG). One package containing ten individually packaged, sterile tracheal tubes of one size only constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances of procedures) and testing (determination by technical mean of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection. Sampling for inspection of all materials shall be conducted with MIL-STD-105 at Level S-2, Acceptable Quality Level (AQL) 1.0%.

Sampling for inspection of the end-item shall be conducted in accordance with MIL-STD-105 at Level S-2, AQL 1.5%. The unit of product for inspection purposes shall be one tracheal tube.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SPECIFICATION SIZE
6515-01-209-0672	TUBE, TRACHEAL, PEDIATRIC, Plastic, Radiopaque, Murphy, 3.5 mm id, 18 cm Long, Disposable, 10s	3.5
6515-01-209-0674	TUBE, TRACHEAL, PEDIATRIC, Plastic, Radiopaque, Murphy, 2.5 mm id, 14 cm Long, Disposable, 10s	2.5
6515-01-252-0038	TUBE, TRACHEAL, PEDIATRIC, Plastic, Radiopaque, Murphy, 5.5 mm id, 26 cm Long, Disposable, 10s	5.5
6515-01-252-2483	TUBE, TRACHEAL, PEDIATRIC, Plastic, Radiopaque, Murphy, 4.5 mm id, 22 cm Long, Disposable, 10s	4.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

NOTE 2: Definitions of appearance of defects listed shall be in accordance with ASTM Designation D883, "Standard Definitions of Terms Relating to Plastics".

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OOS
PHS
FDA-MPQAS

Project No. 6515-4557

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53230A	TUBE, TRACHEAL, Pediatric (Murphy)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include A Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4968 WASHINGTON D C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.