MIL-V-21701B
Amendment 2
21 July 1977
Superseding
Amendment 1
29 September 1970

MILITARY SPECIFICATION

VALVE, CHECK, OXYGEN MASK, INLET

This amendment forms a part of Military Specification MIL-V-21701B, dated 26 March 1969, and is mandatory for use by all Departments and Agencies of the Department of Defense.

Page 1

2.1 SPECIFICATIONS - Military: Delete the following: "MIL-B-22205 Bag Transparent, Flexible, Heat Sealable, for Packaging Applications".

Add the following: "MIL-B-117 Bags, Sleeves and Tubing, Interior Packaging".

Add the following to the bottom of the page:

"Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commanding Officer, Naval Air Engineering Center, Attention: Engineering Specifications and Standards Department, Lakehurst, NJ 08733, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter."

Page 3

3.1 Add the following to the end of the paragraph: "In addition, the retention of the qualification for the valves on the applicable Qualified Products List shall be dependent on periodic verification of continued compliance with the requirements of this specification (see 4.3.2)."

Page 4

4.2 (a) (1) Add the following new paragraph: "(1) Retention of qualification - Retention of qualification consists of periodic verification to determine compliance of the qualified product with the requirements of this specification (see 3.1, 4.3, 4.3.2)"

FSC 1660

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T10:20Z
Check the source to verify that this is the current version before use.

Page 5

4.3.1 Lines 5 and 6 - Delete "Johnsville, Warminster, Pennsylvania 18974, Attention: Director, Aerospace Crew Equipment Department" and substitute "Warminster, Pennsylvania 18974, Attention: Code 4041".

4.3.2 Add the following new paragraph:

"4.3.2 Retention - The retention of qualification shall consist of periodic verification to determine compliance of the qualified valve with the requirements of this specification, the time and method of periodic verification shall be specified by the activity responsible for the Qualified Products List and shall be included in the Notice of Qualification letter."

Page 9

Figure 2 Delete and substitute new Figure 2.

Page 10

Figure 3 Delete and substitute new Figure 3.

Page 11

4.6.4 Line 4, delete "Proof pressure" and substitute "Pressure drop".

Page 14

4.7.2 Line 5, delete "manometer" and substitute "flowmeter".

Page 16

5.1.1 Line 6, delete "MIL-B-22205, Type III" and substitute "MIL-B-117, Type II, Class B, Style 2".

Page 18

6.4 Delete and substitute the following:

"6.4 With respect to products requiring qualification, awards will be made only for products which are at the time set for opening of bids, qualified for inclusion in the applicable Qualified Products List whether or not such products have actually been so listed by that date. The attention of suppliers is called to this requirement,


Source: https://assist.dla.mil -- Downloaded: 2026-10-19T10:20Z
Check the source to verify that this is the current version before use.

and manufacturers are urged to arrange to have the products that they propose to offer to the Federal Government, tested for qualification in order that they may be eligible to be awarded contracts or orders for the products covered by this specification. The activity responsible for the Qualified Products List is the Commander, Naval Air Systems Command, Department of the Navy, Washington, DC 20360; however, authorization for qualification of products shall be obtained from the Commanding Officer, Naval Air Development Center, Warminster, Pennsylvania 18974, (Code 4041). Prior to submission of the samples for qualification inspection, the manufacturer shall submit a request to the Naval Air Development Center (Code 4041) indicating a date on which the samples can be forwarded and also request an authorization number to accompany the samples."

Custodians:
Navy - AS
Army - AV

Preparing Activity:
Navy - AS
(Project No. 1660-0345)


Source: https://assist.dla.mil -- Downloaded: 2026-10-19T10:20Z
Check the source to verify that this is the current version before use.



DIMENSIONS IN INCHES. TOLERANCES: DECIMALS ±.010, ANGLES ±1/2°.

FIGURE 2. TEST FIXTURE (FEMALE END)

Source: https://assist.dla.mil -- Downloaded: 202610-19T10:20Z
Check the source to verify that this is the current version before use.



DIMENSIONS IN INCHES. TOLERANCE: DECIMALS ±.010

FIGURE 3. TEST FIXTURE SECTION WITH PIEZOMETER PRESSURE TAP INCLUDED

Source: https://assist.dla.mil -- Downloaded: 202610-19T10:20Z
Check the source to verify that this is the currentversion before use.