A-A-54467
30 September 1991

COMMERCIAL ITEM DESCRIPTION

SPONGE, SURGICAL, GAUZE (COMPRESSED)

The General Services Administration has authorized the use of this Commercial Item Description in preference to portions of Type I, portions of Classes A and B of Interim Federal Specification DDD-P-0054A (GSA-FSS).

Salient Characteristics:

Shall be nonsterile gauze sponges suitable for use in the treatment of injuries and in surgical procedures.

Sponges shall be of the following types, as specified:

TYPE	FINISHED FOLDED SIZE (INCHES)	TOLERANCE INCH	NUMBER OF PLIES
1	4 by 4	± 3/8	8
2	4 by 8	± 1/2	12
3	2 by 2	± 1/4	12

Sponges shall be fabricated from gauze meeting all requirements of Absorbent Gauze, USP, Type VII, except the filling count may be 8 to 11 threads/inch. The minimum area of the piece of gauze from this the sponge is folded shall be as specified below. Sponges shall be folded in such a manner that the finished sponges shall not have any exposed raw edges or loose threads. Sponges shall be capable of being opened to the sizes specified below without exposing any raw edges or loose threads.

TYPE	OPEN SIZE (INCHES)	MINIMUM AREA OF GAUZE (SQUARE INCHES)
1	4 by 8	125
2	4 by 24	376
3	2 by 6	47

The contractor may, at his option, include a precount feature for each 10 or 20 sponges in each unit package. Any precount feature (if present) shall not deteriorate when autoclaved and any precount feature which is colored shall be colorfast when autoclaved. Sponges shall be free of cuts, holes, tears, foreign matter and ravelings.

AMSC\NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the currentversion before use.

Workmanship. Sponges shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 200 sponges of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one sponge. Sampling shall be Inspection level S-4 with an acceptable quality level (AQL) of 2.5 (percent defective).

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one sponge. Sampling shall be Inspection level S-1 with an AQL of 2.5 (percent defective).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the currentversion before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE
6510-00-782-2698	SPONGE, SURGICAL, GAUZE, Compressed, 4 by 4 inches, 200s	1
6510-00-782-2699	SPONGE, SURGICAL, GAUZE, Compressed, 4 by 8 inches, 200s	2
6510-00-782-2700	SPONGE, SURGICAL, GAUZE, Compressed, 2 by 2 inches, 200s	3

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC 05205

Project No. 6510-0783

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54467	30 SEP 91

3. DOCUMENT TITLE
SPONGE, SURGICAL, GAUZE (COMPRESSED)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/29

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the currentversion before use.