A-A-54468
30 September 1991

COMMERCIAL ITEM DESCRIPTION

ROLL, ABSORBENT

The General Services Administration has authorized the use of this Commercial Item Description in preference to Type I of Roll, Absorbent, Cotton and Roll, Cellulose Specification GG-R-563B.

Salient Characteristics:

Shall be sterile absorbent cotton rolls suitable for use in dental procedures. The rolls shall be sterile when tested in accordance with the USP. The rolls shall be of the following sizes and fifty of the finished rolls shall have the following minimum average weights.

SIZE	DIMENSION (INCHES)	MINIMUM AVERAGE WEIGHT OF 50 ROLLS (GRAMS)
1	3/8 Diameter By 6 Long	70
2	1/2 Diameter By 1 1/2 Long	25

The rolls shall be made from cotton that meets the requirements of Purified Cotton, USP, except for fiber length and except that it shall retain not less than 20 times its weight of water.

The cotton shall be formed into non-collapsible rolls of uniform firmness and consistency. The outer surface of the finished rolls shall be starched or shall be braided with a soft synthetic yarn, or shall be coated with methylcellulose in such a manner as to prevent adhesion or damage to oral mucous membranes and still be capable of uniform absorption rate. The rolls shall be readily adaptable to mouth contour and shall have a high degree of absorbency. The rolls shall be well formed and shall be free from dirt, lumps, foreign matter and other impurities.

AMSC\NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Workmanship. The absorbent rolls shall be free from defects which detract from their appearance or impair their serviceability.

Unit.

For Size 1 - Box (BX). Fifty absorbent cotton rolls shall be suitably bagged or wrapped and sealed to maintain sterility of contents unless packaged is damaged or opened. One box containing 10 packages of 50 rolls each (500 rolls), as specified, constitutes one unit. As an alternate, 50 absorbent cotton rolls shall be suitably banded in a bundle. Five bundles of 50 rolls each (250 rolls) shall be suitably bagged and sealed to maintain sterility of contents unless package is damaged or opened. One box containing two packages of 250 rolls each (500 rolls), as specified, constitutes one unit.

For Size 2 - Box (BX). Individual absorbent cotton rolls shall be suitably bagged, wrapped or bundled in groups of not less than 25 rolls nor more than 125 rolls. Each group shall be suitably bagged or wrapped and sealed to maintain sterility of contents unless package is damaged or opened. One box containing a total of 2000 bagged or wrapped rolls, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one roll. Sampling shall be Inspection level S-4 with an acceptable quality level (AQL) of 2.5 (percent defective).

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shal be one roll. Sampling shall be Inspection level S-1 with an AQL of 2.5 (percent defective).

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-518-1900	ROLL, ABSORBENT, COTTON, 3/8 By 6 Inches, 500s	1
6510-01-073-1331	ROLL, ABSORBENT, COTTON, 1/2 By 1 1/2 Inches, 2000s	2

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC 05086

Project No. 6510-0778

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54468	2. DOCUMENT DATE (YYMMDD) 30 SEP 91

3. DOCUMENT TITLE
ROLL, ABSORBENT

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete. 198/290

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.