A-A-54519
12 November 1991

COMMERCIAL ITEM DESCRIPTION

THERMOMETER, SELF-INDICATING, LIQUID IN GLASS (Blood Bank)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a low temperature range thermometer for use with blood bank refrigerators.

Salient Characteristics:

Shall be mercury-filled glass stemmed thermometers that are calibrated for total immersion. Shall be of the following type thermometers:

Type I. A thermometer that is marked by the manufacturer with a serialized identification number.

Type II. A thermometer that is not required to be marked with a serialized identification number, but can be supplied with a serialized identification number.

Shall be suitable for use with blood bank refrigerators in checking that the actual chamber temperature is within a required range of +4 degrees Celsius (C) to +6 degrees (C). Shall be suitable for placement in a simulated blood collection bag that is used in a refrigerator chamber to monitor the actual blood temperature.

Shall be graduated in one (1) degree (C) divisions for a range that includes -5 to +20 degrees (C). The graduations shall have an accuracy of +/- 1 degree (C). Etched arrows shall be marked with graduations to indicate a critical range of +4 to +6 degrees (C).

Shall have a glass stem incorporating legible and permanent markings that include the required graduations, Arabic numerals for at least the 0 and 10 degrees (C) divisions, the etched arrows indicating the critical temperature range, and the manufacturer's name or trademark. The type I thermometer shall also be marked to include a serialized identification number.

AMSC N/A FSC 6685

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Shall have an overall length of 152 millimeters (mm) +/-5 mm. The top of the glass stem shall have a button top type finish that is perpendicular to the axial length of the thermometer. The button top diameter shall be greater than that of the body of the stem.

Workmanship. The thermometer shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One thermometer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall also be conducted in accordance with commercial practice. In addition each thermometer shall be checked for graduation accuracy against standard thermometer(s) whose calibration is traceable to the National Institute of Standardsand Technology (NIST). The graduation markings selected for testing shall include the +4 and +6 degree (C) temperature divisions. A thermometer that does not conform to the +/- 1 degree (C) accuracy requirement at any point on the required scale markings shall be cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 20260-26T09:35Z
Check the source to verify that this is the current version before use.

degree of bias, qualified for inclusion in the appropriate

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SPECIFICATION DESIGNATION
6685-01-167-6662	THERMOMETER, SELF-INDICATING, LIQUID IN GLASS.	Type II

This thermometer is intended to comply with the requirements of U. S. Food and Drug Administration (FDA) and American Association Blood Banks, for monitoring the temperature of blood in blood storage refrigerators.

Ordering data. (Type thermometer, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DLA-DM

Project Number: 6685-0873

Location: ENABLE/THERMCID/S31/2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Connection with specific acquisition functions should be obtained fro-

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

RECOMMENDED A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54519	12 November 1991

3. DOCUMENT TITLE

THERMOMETER, SELF-INDICATING, LIQUID IN GLASS (Blood Bank)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)

b. ORGANIZATION

c. ADDRESS (Include Zip Code)

d. TELEPHONE (Include Area Code)
(1) Commercial
(2) AUTOVON
(If applicable)

7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME

Defense Personnel Support Center

b. TELEPHONE (Include Area Code)
(1) Commercial: (215) 737-2870
(2) AUTOVON: 444-2870

c. ADDRESS (Include Zip Code)
ATTN: DPSC-MSE
2800 S. 20th Street
Philadelphia, PA 19101

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.