A-A-54528
23 November 1991

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a sterile, disposable, surgical drape pack for creating a sterile field for abdominal surgical procedures.

Salient Characteristics:

The pack shall be a sterile, disposable, surgical drape pack fabricated of nonwoven material for creating a sterile field specifically for abdominal surgical procedures. The pack shall consist of at least the following components:

* One overwrap/table cover, approximately 44 by 90 inches
* One Mayo stand cover with pad, approximately 23 by 55 inches
* One scrub nurse gown with towel, overwrapped
* Two surgical gowns with towels
* Four utility drapes with tapes, approximately 14 by 26 inches
* One suture bag
* One laparotomy T-sheet with a bactericidal fabric reinforcement, an instrument pad, adjustable adhesive fenestration, 3 feet by 12 inches, cord holders and armboard covers

The overwrap/table cover shall be fan-folded for ease in opening on the surgical back table. The cover shall be impervious to fluids and moisture.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

The Mayo stand and tray cover shall be made of a plastic material at least 2.75 mils thick and shall be impervious to fluids. The cover shall be fan-folded with a 6 inch cuff on the distal end for sterile draping of the Mayo stand. The pad shall be large enough to cover the surface of the Mayo tray.

The three (3) gowns shall be large or extra large size. All gowns shall:

* Be soft, comfortable, pliable, quiet and memory free
* Have high breathability facilitating comfort for wearing for long period of time
* Be fluid repellent
* Have stockinette cuffs that shall conform to the wrists
* Be the wrap-around type to completely cover the back of the individual gowned
* Have a tie closure that shall not rip or separate when tying. The gown shall be designed to minimize the possibility of contamination when tying.
* One gown with towel shall be in an overwrap for the scrub nurse

The four (4) utility drapes shall be water repellent and have an adhesive strip on the back for skin adherence.

The pack shall have a T-shaped laparotomy sheet, approximately 123 inches long with an adjustable 3 feet by 12 inch fenestration. The reinforcement area shall be impervious to fluids yet be highly absorbent, eliminating strike-through while preventing pooling of fluids. The remainder of the drape shall be fluid/moisture resistant. The sheet shall have ESU cord/suction tubing holders and shall contain an instrument pad as part of the drape. The sheet shall be cuffed and fan-folded to permit aseptic draping of the operative area.

The pack shall have a nonwoven fabric or plastic suture bag.

All nonwoven material shall be fabricated of a soft material with minimal memory for good drapeability. The material shall resist tearing, pilling and linting, and shall be fluid/water repellent.

If the pack is opened and not used, it shall be capable of re-sterilization by ethylene oxide.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

The packs shall be sterile and individually wrapped in a suitable tear-resistant, waterproof, heat sealed package which shall allow aseptic handling and maintain sterility of the contents until opened or damaged. Items shall be packed in sequence of use.

Sterility of the packs shall be in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001.

All pack components shall be clearly listed on the exterior wrapper which shall also serve as a dust cover.

Workmanship. The surgical pack shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing four individually packaged sterile laparotomy packs, as specified, constitutes one unit. The individual package shall be a suitable container capable of maintaining sterility of contents unless package is damaged or opened. Items shall be packed in sequence of use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one package of four laparatomy packs.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-I. The unit of product for test purposes shall be one package of four laparotomy packs.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sterility test. The surgical packs shall be sterile in accordance with the requirements of the Food and Drug Administration.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6530-01-145-6263

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

A-A-54528

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA-MPQAS

Project Number: 6530-2300

Location: PACKMC15.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54528	2. DOCUMENT DATE (YYMMDD) 23 November 1991

3. DOCUMENT TITLE
SURGICAL PACK, DISPOSABLE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.