A-A-54529
23 November 1991

COMMERCIAL ITEM DESCRIPTION

ULTRASONIC CLEANING SYSTEM, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an ultrasonic cleaning system for use in cleaning dental instruments.

Salient Characteristics:

The unit shall be an ultrasonic cleaning system for use in cleaning dental instruments. The cleaning system shall deliver ultrasonic power with a nominal main frequency not less than 20 kHz. The power density shall be not less than 40 watts per gallon.

The system shall be a benchtop unit with nominal outside dimensions of 13.5 inches by 12 inches by 11 inches.

The tank shall have a capacity of 2.5 gallons. The system shall be capable of accommodating three (3) Instrument Management System (IMS) instrument cassettes, each having a size of 8 inches by 4.5 inches by 1.25 inches.

The system shall be supplied with a tank lid cover and cassette rack.

The system shall have an automatic timer, which may be digital or analog, a thermostatically controlled heater and a rear-mounted drain with valve.

The system shall have an internal power line filter to eliminate electrical interference with other nearby electrical equipment. This power line filter shall not generate more than 500 microamperes leakage current.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

The system shall comply with the Code of Federal Regulations (CFR) 47, Section 18, and applicable UL Standards, which shall be evidenced by UL listing, CSA approval, City of Los Angeles approval, or testing by an independent laboratory.

The ultrasonic cleaning system shall be capable of operating over the full tolerance range of one or more of the following line voltages and frequencies as specified by the customer requisition:

* 115 VAC plus or minus 10 percent, 50 Hz plus or minus 2 Hz
* 115 VAC plus or minus 10 percent, 60 Hz plus or minus 2 Hz
* 230 VAC plus or minus 10 percent, 50 Hz plus or minus 2 Hz
* 230 VAC plus or minus 10 percent, 60 Hz plus or minus 2 Hz

Dual voltage and hertz on all units are desirable but not mandatory.

A three-wire power cord and plug, which has been approved by a nationally recognized testing laboratory as suitable for hospital environments, shall be supplied.

Operational and Service Data Manuals shall be furnished as follows:

a. Information Content:

The contractor shall furnish manuals, handbooks, and/or brochures containing complete operation, installation, and service/maintenance instructions (including pictures or illustrations, as necessary) with complete schematics and wiring diagrams. All manuals, handbooks, and/or brochures shall be written in the English language, and all schematics and wiring diagrams shall use American electrical and electronics symbols. The manuals shall include electrical data and connection diagrams for all applicable utilities. The instructions shall also contain a complete list of all replaceable parts showing part number, name, and quantity required. All service notes, service memos, etc. pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included. These instructions shall be, as a minimum, those furnished to service engineers (or servicemen) who normally install and service the equipment for the company or distributor. When the system being procured includes, or will operate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

ancillary equipment, the service data shall include complete instructions and drawings that show the interfacing of all system components.

b. Quantities, Distribution and Schedules:

Two complete sets of the above information shall be packaged with each unit of the required equipment. Where multiple quantities of the same unit are to be delivered to the same ultimate consignee, only a total of two service manuals need to be supplied to each consignee. However, two operational manuals still need to be supplied with each unit.

Workmanship. The ultrasonic cleaning system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One ultrasonic cleaning system, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:35Z
Check the source to verify that this is the current version before use.

properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one ultrasonic cleaning system.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one ultrasonic cleaning system.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

The ultrasonic cleaning system shall be tested to determine proper operation at a selected supply power. The unit shall be tested to verify that the ultrasonic power produced has a frequency greater than 20 kHz and a power density greater than 40 watts per gallon. This test shall be conducted at each supply power specified.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:35Z
Check the source to verify that this is the current version before use.

Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:35Z
Check the source to verify that this is the current version before use.

A-A-54529

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6520-2481

Location: CLEANCID.16D

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54529	2. DOCUMENT DATE (YYMMDD) 23 November 1991

3. DOCUMENT TITLE
ULTRASONIC CLEANING SYSTEM, DENTAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.