A-A-54530
25 November 1991

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a sterile, disposable surgical pack for use in creating a sterile field for extremities.

Salient characteristics:

The sterile, disposable, surgical drape pack shall consist of:

One (1) overwrap/table cover, approximately 50 by 90 inches.

One (1) Mayo stand cover.

Four (4) drape towels with adhesive tapes, approximately 17 by 26 inches.

One (1) suture bag.

One (1) plastic sheet, fan-folded, approximately 44 by 60 inches.

One (1) extremity sheet with a bactericidal absorbent reinforcement, approximately 87 by 128 inches with a 2.5 inch circular elastic fenestration and ESU cord/suction tubing holders.

The overwrap/table cover shall be reinforced with a polyethylene back. Shall be fan-folded for ease in opening on the surgical back table. The overwrap shall be impervious to fluids and moisture.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Sterility test. The surgical packs shall be tested for sterility in accordance with the requirements of the Food and Drug Administration.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:35Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6530-01-184-2404	SURGICAL PACK, DISPOSABLE, 5s, Sterile.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6530-2301

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:35Z
Check the source to verify that this is the current version before use.

The Mayo tray and stand cover shall be made of a plastic material at least 2.75 mils thick. The tray portion shall be reinforced with a non-woven material measuring 28 inches by 20 inches and shall be impervious to fluids. The cover shall be fan-folded with a 6 inch cuff on the distal end for sterile draping of the Mayo stand.

The four drape towels shall be of a soft non-woven fabric with an adhesive strip the length of the towel for adherence to the skin.

The suture bag shall be made of plastic or a non-woven fabric.

The polyethylene sheet shall be used as a water/fluid barrier. It shall go under the extremity before the drapes are applied.

The extremity sheet shall be approximately 87 by 128 inches long. Shall have a bactericidal fabric reinforcement around the fenestration. The reinforcement area shall measure 27 inches by 46 inches and shall be impervious to fluids yet be highly absorbent, eliminating strike through while preventing pooling of fluids. The remainder of the drape shall be fluid/moisture resistant. Shall be fan-folded to permit aseptic draping of the operative area.

It shall be possible to re-sterilize the pack by ethylene oxide, should the packs be opened and not used.

All non-woven material shall be fabricated of a soft material with a minimal memory for good drapeability. Shall resist tearing, pilling and linting.

The pack shall be sterile and individually wrapped in a 4 mil poly bag. The bag shall be a tear resistant, waterproof, heat sealed package which shall allow aseptic handling.

The manufacturer's lot number shall be heat-sealed into the poly bag for traceability purposes.

All pack components shall be clearly listed on the exterior wrapper which shall also serve as a dust cover.

Workmanship. The surgical packs shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:35Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing five (5) individually packaged sterile surgical packs, as specified, constitutes one unit. The individual package shall be a sealed peel-open type container which allows aseptic handling and which shall be capable of maintaining sterility of contents unless the package is damaged or opened. Items shall be packed in sequence of use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item. Sample masks shall be examined for consumer type defects: holes, cuts, tears, and loose or missing components.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical pack.

Sampling for tests. Sampling for tests including dimensional tests, shall be conducted with and AQL of 2.5 (percent defective) and an inspection level of 1.0. The unit of product for test purposes shall be one surgical pack.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54530	2. DOCUMENT DATE (YYMMDD) 25 November 1991

3. DOCUMENT TITLE
SURGICAL PACK, DISPOSABLE

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME *(Last, First, Middle Initial)*	b. ORGANIZATION	
c. ADDRESS *(Include Zip Code)*	d. TELEPHONE *(Include Area Code)* (1) Commercial (2) AUTOVON *(If applicable)*	7. DATE SUBMITTED *(YYMMDD)*

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center	b. TELEPHONE *(Include Area Code)* (1) Commercial (215) 737-2870 (2) AUTOVON 444-2870
c. ADDRESS *(Include Zip Code)* ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145.	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil/Downloads/26104/26109357
Check the source to verify that this is the current version before use.