A-A-54541
6 December 1991

COMMERCIAL ITEM DESCRIPTION

SPREADER, PLASTIC CAST (Stryker)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a plastic cast spreader of the Stryker design suitable for separating saw-cut plastic casts with single-hauled operation.

Salient Characteristics:

General. The spreader shall have 2 jaws on each blade. Each blade shall be serrated to the outside (see drawing). Each handle shall have a safety stop. The handles shall be spring-loaded with the spreader halves held together with 2 screws. The spreader weight shall not exceed 14 ounces.

Material. Material of the spreader shall be either Class 4 martensitic stainless steel in accordance with ASTM Designation F899, or annealed, chromium-over-nickel plated steel alloy AISI No. 6150H or 8650H. The spring, screws, and pivot pin shall be fabricated from Class 3 austenitic stainless steel in accordance with ASTM Designation F899.

Style, design and dimensions. The spreader shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24292 for style, design and dimensions.

Hardness. The jaws shall have a Rockwell Hardness C scale (HRC) of 50 to 55 and the pivot pin shall have a Rockwell Hardness C scale (HRC) 30 to 35 when tested in accordance with ASTM Designation E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM Designation E140.

Finish. The spreader shall have a mirror, satin or matte finish. The matte finish shall be exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure. Plated surfaces shall be bright and smooth.

Passivation. The spreader shall be properly passivated.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Corrosion resistance. Parts fabricated from Class 4 steel shall show no signs of corrosion or plating of copper when tested in accordance with ASTM Designation F1089, and Class 3 and AISI carbon steel 6150H or 8650H parts shall show no signs of corrosion when boil tested in accordance with ASTM Designation F1089.

Performance. The jaws shall align as shown on the drawing in the closed position. Jaw separation shall be uniform in depth and spacing. Following the boil test, the spreader shall be opened and closed 5 times using one hand. Action shall be without binding and the spring shall not slip out of its cavity.

Identification marking and labeling. Spreaders shall be legibly marked with the contractor's name or trademark and the word "stainless" (if it is fabricated from stainless steel). Foreign manufactured spreaders shall bear the name of the country of origin.

Workmanship. The spreader shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One spreader, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:35Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one spreader. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one spreader. Sampling shall be inspection level shall be S-2 with and AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its Item Identification is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-935-9880	SPREADER, PLASTER, CAST, Stryker

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:35Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

PGC 46624

Project No. 6515-4944
location: AA54541/S17
disk 4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54541	2. DOCUMENT DATE (YYMMDD): 6 December 1991

3. DOCUMENT TITLE: SPREADER, PLASTIC CAST (Stryker)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: (215)7372870	(2) AUTOVON: 444-2870
ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.