A-A-54550
20 December 1991

# COMMERCIAL ITEM DESCRIPTION

## TUBING, ASSEMBLY, BAG, ASEPTIC SOLUTION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, disposable, rigid plastic tube suitable for aseptic decanting of I.V. solutions from flexible containers.

Salient Characteristics:

General. The tubing assembly shall be approximately 8 inches long with a spike at one end and having a slight bend at the other end.

Style, design, and dimensions. The end adjacent the spike shall have a collar to prevent entry of the piercing spike into the bag beyond a certain length. The spike and opposite open end shall be covered with removable plastic caps.

The devices shall be in accordance with Figure 1.

Material. The tubing assembly shall be fabricated from acrylic plastic XT polymer 250 or other suitable plastic.

Sterility. The tubing assembly and internal surfaces of the individual package shall be in accordance with USP or AAMI guidelines. Sterility assurance level shall be 0.000001.

Shelf life. The assembly shall have a 36 month shelf life with at least 30 months of shelf life remaining at the time of delivery to the Government.

Workmanship. The tubing assembly shall be free from defects which detract from its appearance or impair its serviceability. The device shall be free from flash, sharp edges, scuff marks. distortion or foreign material.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing 25 individually packaged sterile tubing assembles, as specified, constitutes one unit. Each assembly shall be packaged in a suitable container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one tubing assembly. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one tubing assembly. Sampling shall be inspection level S-2 with an AQL of 1.0.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-173-7030 | TUBING ASSEMBLY, BAG, ASEPTIC SOLUTION |

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:


VA-OSS
PHS
FDA-MPQAS

PGC S0220

Product No. 6515-5305
Location: A-A-54550\Nish disk 1



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.




| UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ON | |
|---|---|
| FRACTIONS | DECIMALS ANGLES |
| ± 1/32 | ± 3° |

A-A-54550

NSN 6515-01-173-7030

FIGURE 1 TUBING ASSEMBLY, BAG, ASEPTIC SOLUTION

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and r letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarificat requirements on current contracts Comments submitted on this form do not constitute or imply authoriza waive any portion of the referenced document(s) or to amend contractual requirements

| I RECOMMEND A CHANGE: | 1 DOCUMENT NUMBER | 2 DOCUMENT DATE (YYMMDD) |
|---|---|---|
| | A-A-54550 | 20 December 1991 |

3 DOCUMENT TITLE

TUBING, ASSEMBLY, BAG, ASEPTIC SOLUTION

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

| a NAME (Last, First, Middle Initial) | b ORGANIZATION |
|---|---|
| c ADDRESS (Include Zip Code) | d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) |
| | 7 DATE SUBMITTED (YYMMDD) |

8 PREPARING ACTIVITY

| a NAME | b TELEPHONE (Include Area Code) (1) Commercial | (2) AUTOVON |
|---|---|---|
| Defense Personnel Support Center ATTN: DPSC-MSE | (215) 737-2870 | 444-2870 |
| c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-346 Telephone (703) 756-2340 AUTOVON 289-2340 | |

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.