A-A-54554
17 January 1992

# COMMERCIAL ITEM DESCRIPTION

## SAW, FINGER RING

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-36635 which is cancelled.

This Commercial Item Description covers a manually operable saw suitable for cutting metal finger rings.

Salient characteristics:

Material. The blade of the saw shall be fabricated from unplated high carbon steel. The bearing washer and hub shall be fabricated from martensitic corrosion-resisting steel (the martensitic corrosion-resisting steel alloy for hub component shall have a manganese content not exceeding 1.80 percent). The handle, tip, and lever shall be fabricated from austenitic corrosion-resisting steel, or martensitic steel, or chromium plated over nickel plated copper alloy. The thumbpiece shall be fabricated from a corrosion-resisting steel or from a combination of molded reinforced nylon and austenitic corrosion-resisting steel. As an alternate, the handle shall be fabricated from molded reinforced nylon. All other fastening devices such as screws, pins or rivets shall be fabricated from austenitic or martensitic corrosion-resisting steel. All corrosion-resisting steel shall be in accordance with ASTM F899.

Hardness. The Rockwell hardness of the blade shall be 73 to 79 on the 30-N scale. The Rockwell hardness of the hub and washer shall be in C scale (HRC) of 38 to 48 when fabricated from martensitic steel, and in B scale 75-95 when fabricated from austenitic steel when tested in accordance with ASTM E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM E140.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The finger ring saw shall be designed to manually cut metal finger rings from the human hand without damage to patient's fingers. The assembly shall consist of a blade that is accurately fitted to a bearing hub and bearing washer that is secured to the handle properly by means of a thumbpiece. The hub shall be designed to form a tight to interference fit (maximum 0.002 inch) with the blade. The thumbpiece shall be readily removed from the handle and bearing hub for easy replacement of the blade. The lever shall be removable, secured to the handle and designed to slip under a finger ring for cutting of the finger rings by manually turning the thumbpiece. The lever slot shall be in alignment with the blade, without the blade butting, jamming or hitting the lever when the lever is moved toward the blade. The washer, blade and hub components shall be readily assembled to each other in any order (i.e. both blade surfaces and both washer surfaces shall be capable of being placed onto the hub.) For saws of style 1 design, the blade shall be prevented from cutting through the lever by having the lever tip coming to a stop at the saw hub and washer. For saws of style 2 design, the lever shall be capable of restraining the blade for blade replacement, by manually turning the thumbpiece in a counter-clockwise direction. The blade shall possess well defined teeth, consistent depth and be in accordance with good manufacturing practice. The finger ring saw assembly, including the blade shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 1820 for style 1 or style 2 design.

Finish. The finish of the finger ring saw assembly shall be smooth and uniform, free from cracks, burrs, sharp edges (except for blade teeth), pits, scale and corrosion.

Passivation. All components fabricated from austenitic or martensitic steel shall be properly passivated.

Corrosion resistance. The plating and corrosion requirements for components fabricated from copper alloy shall be in accordance with ASTM Standard B117, except that the corrosion test shall be the boil test only in accordance with ASTM F1089. Except for the blade, components fabricated from carbon steel shall comply with corrosion tests in accordance with ASTM F1089, except that the boil test only shall apply. Components fabricated from austenitic or martensitic steel shall show no corrosion or plating of copper when tested in accordance with ASTM F1089. Components fabricated from austenitic steel shall show no corrosion when subjected to boil test in accordance with ASTM F1089.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Performance tests. A sufficient straight length of 1/16 inch minimum diameter, brass wire or rod of quarter-hard minimum temper shall be used as the test specimen. The test specimen shall be free from cracks, pits or other evidence of unsuitability. The test specimen shall be securely placed between the lever tip and blade. The saw thumbpiece shall be turned using nominal technique (without excessive pressure) in clockwise manner. To minimize slippage of the test specimen, the test specimen may be appropriately restrained.

Performance. When tested as specified above, the saw shall cut completely through the test specimen within two complete revolutions without breaking the blade teeth.

Identification marking and labeling. The finger ring saw shall be legibly marked with the contractor's name or trademark and the word "stainless". Foreign manufactured instruments shall bear the name of the country of origin.

Workmanship. The finger ring saw assemblies and blades shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One finger ring saw, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:35Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one finger ring saw. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one finger ring saw. Sampling shall be Inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:35Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

NOTE: The following National Stock Numbers are covered by this document:

| NSN | Item Identification |
|---|---|
| 6515-00-363-4100 | SAW, FINGER RING |
| 6515-00-363-4150 | BLADE, FINGER RING SAW |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

A-A-54554

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-4797



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

| I RECOMMEND A CHANGE: | 1 DOCUMENT NUMBER A-A-54554 | 2 DOCUMENT DATE (YYMMDD) 17 January 1992 |
|---|---|---|

3 DOCUMENT TITLE SAW, FINGER RING

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

| a NAME (Last, First, Middle Initial) | b ORGANIZATION |
|---|---|
| c ADDRESS (Include Zip Code) | d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e DATE SUBMITTED (YYMMDD) |

8 PREPARING ACTIVITY

| a NAME Defense Personnel Support Center | b TELEPHONE (Include Area Code) (1) Commercial (215) 737-2870 (2) AUTOVON 444-2870 |
|---|---|
| c ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.