DD-C-00791d
October 24, 1963
Interim revision of
Fed. Spec. DD-C-791c
August 14, 1963

INTERIM FEDERAL SPECIFICATION

CUP, GLASS; PIN, CLIP, AND SPONGE

This Interim Federal Specification was prepared by the General Services Administration, Federal Supply Service, Standardization Division, Washington, D. C., 20407, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward any recommendations for changes to the preparing activity. The General Services Administration has authorized Federal agencies to use this Interim Federal Specification as a valid exception to Federal Specification DD-C-791c, dated August 14, 1963.

1. SCOPE AND CLASSIFICATION

1.1 Scope. Cups covered by this specification are intended for use in containing pins, clips, and sponges.

1.2 Classification.

1.2.1 Type and grade. Pin, clip, and sponge cups covered by this specification shall be of one type and one grade.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification to the extent specified herein:

Federal Specifications:

PPP-B-576 - Box, Wood, Cleated, Veneer, Paper Overlaid.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-591 - Boxes, Fiberboard, Wood-Cleated.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Box, Fiberboard.
PPP-T-97 - Tape; Pressure-Sensitive Adhesive, Filament Reinforced.

Federal Standards:

Fed. Std. No. 102 - Preservation, Packaging, and Packing Levels.
Fed. Std. No. 123 - Marking for Domestic Shipment (Civilian Agencies).

FOR INSTRUCTIONS ON OBTAINING COPIES OF FEDERAL SPECIFICATIONS AND STANDARDS, SEE LAST PAGE.

Military Specification:

MIL-L-10547 - Liners, Case, Waterproof.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 7510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. The glass used in the manufacture of the cups shall be clear and have such quality and strength that the cups will not crack or break when tested as specified in 4.4.

3.2 Dimensions. The over-all diameter shall be 3-3/4 inches ± 1/4 inch. The height shall be 1-5/8 inches ± 1/8 inch.

3.3 Design. The cups shall be circular in cross section. The internal walls shall be slightly concave. All edges shall be rounded. The base shall rest firmly on any flat surface and shall be so designed that the cup will not tilt when in use.

3.4 Workmanship. Cups shall be free from foreign material, sandholes, bubbles, chips and other defects which may affect their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Supplier's responsibility. Unless otherwise specified (see 6.1), the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.2 Sampling.

4.2.1 Size of lot. All cups presented for delivery at one time shall be considered a lot for the purpose of examination and tests.

4.2.2 Sampling for visual and dimensional examination and for tests. A random sample of cups shall be taken from each lot in accordance with MIL-STD-105, inspection level L-6. Lot acceptance or rejection shall be based on an AQL of 6.5 percent defective (see 4.3.1).

4.2.3 Sampling for examination of preparation for delivery. Sample shipping containers of packaged and packed cups shall be taken from each lot in accordance with MIL-STD-105, Inspection Level L-5, with an Acceptable Quality Level (AQL) of 4.0 percent defective. The sample unit shall be one shipping container.

4.3 Examination.

4.3.1 Visual and dimensional examination. Sample cups, selected in accordance with 4.2.2, shall be examined for the defects listed in table I. Any cup having one or more defects shall be considered a defective cup. If the number of defective cups exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

TABLE I. - Classification of defects

Defects (Major):
Glass not clear
Cup tilts when in use.
Base does not rest firmly on flat surface .
Edges not rounded.
Surface not slightly concave.
Dimensions not as specified.
Workmanship: Any holes, bubbles, chips, cracks, or rough edges.

4.3.2 Examination of preparation for delivery. Sample shipping containers, selected in accordance with 4.2.3, shall be examined for compliance with packaging, packing, and marking requirements in section 5 of this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

4.4 Test. The same samples, selected in accordance with 4.2.2 shall be subjected to the following test. Any cup failing this test shall be considered a defective cup. If the number of defective cups exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

4.4.1 Dropping test. The cups shall be dropped edgewise three times from a height of 30 inches onto a wooden floor. Examine for compliance with the requirements of 3.1.

5. PREPARATION FOR DELIVERY

(For Civil agency procurement, the definitions and applications of the various levels of packaging and packing specified hereinafter shall be in accordance with Fed. Std. No. 102.)

5.1 Packaging. Cups shall be packaged level A, B, or C, as specified in the contract or order (see 6.1).

5.1.1 Level A. Twelve cups shall be packaged in a snug-fitting box conforming to PPP-B-636, type I or II, class 2, style optional. Each cup shall be separated by individual corrugated fiberboard cells in a manner which will restrict movement. Fiberboard for box and separators shall be not less than 175 p.s.i. The box shall be closed and sealed in accordance with the appendix to the box specification.

5.1.2 Level B. Twelve cups shall be packaged in a snug-fitting box conforming to PPP-B-636, type, class, and style optional. Each cup shall be separated by individual corrugated fiberboard cells in a manner which will restrict movement. Fiberboard for box and separators shall be not less than 175 p.s.i. Closure shall be in accordance with the appendix to the box specification.

5.1.3 Level C. Cups shall be packaged in accordance with the supplier's commercial practice.

5.2 Packing. Cups, packaged level A, B, or C, shall be packed level A, B, or C, as specified in the contract or order (see 6.1).

5.2.1 Level A. Six unit packages (72 cups) shall be packed in overseas-type or class boxes conforming to any of the following specifications:

PPP-B-585, class 3 - Boxes, Wood, Wire-bound.

PPP-B-591 - Boxes, Fiberboard, Wood-Cleated.

PPP-B-601 - Boxes, Wood, Cleated-Plywood.

PPP-B-621 - Boxes, Wood, Nailed and Lock Corner.

PPP-B-636, V3c or V3s - Box, Fiberboard, class 2.

PPP-B-576 - Box, Wood, Cleated, Veneer, Paper Overlaid.

Box closures and strapping shall conform to the applicable box specification and the appendix thereto. In lieu of steel strapping, fiber boxes may be reinforced with pressure sensitive filament tape conforming to PPP-T-97, type IV, applied in accordance with the appendix thereto. Boxes shall be provided with case liners conforming to MIL-L-10547, and sealed in accordance with the appendix thereto.

5.2.2 Level B. Six units, packaged as specified in 5.1.2, shall be packed in a box conforming to PPP-B-636, type I or II, class 1, style optional, minimum bursting strength 275 p.s.i. Sealing and closure shall be in accordance with the box specification.

5.2.3 Level C. Cups shall be packed in containers of the type, size, and kind commonly used for the purpose, in a manner that will insure acceptance by common carrier and safe delivery at destination. Shipping containers shall comply with the regulations of carriers as applicable to the mode of transportation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Civil agencies. In addition to markings specified in the contract or order, the packages and shipping containers shall be marked in accordance with Fed. Std. No. 123, including specification classification, such as type, size, etc.

5.3.2 Military agencies. In addition to any special marking specified in the contract or order, shipments shall be marked in accordance with MIL-STD-129.

5.4 Standard pack. The standard pack requirements for the items covered by this specification, including marking, shall be as specified in 5.1.2 (level B), 5.2.2 (level B), and 5.3.1 (see 6.1.1).

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein and procurement documents should specify the following:

(a) Title, number, and date of this specification.

(b) Inspection responsibility if other than as specified herein; and administrative provisions for inspection records, when applicable (4.1).

(c) Level of packaging and packing required (5.1 and 5.2).

6.1.1 Standard pack (civil agencies). The standard pack requirements, including marking, in 5.1.2 (level B), 5.2.2 (level B), and 5.3.1 are intended for use in procurement of Stores Stock replenishments. Procuring officers should use the standard pack provisions when it is known that the material will be shipped from a supplier to a domestic warehouse, supply depot, or intermediate storage point for temporary storage, subsequent issue, and/or shipment to eventual user.

6.2 Transportation description. Transportation descriptions and minimum weights applicable to this commodity are:

Rail:

Glassware, not otherwise indexed by name.

Carload minimum weight 20,000 pounds, subject to Rule 34, Uniform Freight Classification.

Motor:

Glassware, not otherwise indexed.

Truckload minimum weight 24,000 pounds, subject to Rule 115, National Motor Freight Classification.

6.3 Federal specifications do not include all types, classes, grades, etc., of the commodities indicated by the titles of the specifications, or which are commercially available, but are intended to cover the types, etc., which are suitable for Federal Government requirements.

6.4 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

NOTICE. (See Last Page).

GSA WASH DC 64-6084

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

ADDENDUM

INSTRUCTIONS ON OBTAINING COPIES OF FEDERAL SPECIFICATIONS AND STANDARDS

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Washington, D. C., Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, and Auburn, Wash.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards and Handbooks from established distribution points in their agencies.)

NOTICE

When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE NAVY

POSTAGE AND FEES PAID
NAVY DEPARTMENT

OFFICIAL BUSINESS

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

ORGANIZATION *(Of submitter)*

CITY AND STATE

CONTRACT NO.

QUANTITY OF ITEMS PROCURED

DOLLAR AMOUNT
$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*

DATE

DD FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE PLATE NO. 15419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.