FEDERAL SPECIFICATION

TUBES, ANTIGEN-DILUTION, GLASS (KAHN)

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers antigen-dilution test tubes suitable for use in the Kahn precipitation test for syphilis.

1.2 Classification.

1.2.1 *Type.*—Antigen-dilution test tubes shall be of but one type.

2. APPLICABLE SPECIFICATIONS

2.1 Specifications. — The following Federal Specifications, of the issues in effect on date of invitation for bids, form a part of this specification:

DD–G–541—Glass (Laboratory).

DD–G–626—Glassware, Laboratory; Methods for Sampling and Testing.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards, and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

2.2 Specifications and standard applicable only to individual departments are listed in section 7.

3. REQUIREMENTS

3.1 Material. — Tubes shall be fabricated from glass conforming to type I or II of Federal Specification DD–G–541.

3.2 Style and design.—Tubes shall be fabricated from glass tubing having an outside diameter of not less than 16½ nor greater than 17¼ mm., and an over-all length of 55 plus or minus 2 mm. The tube shall have a wall thickness of not less than 1.0 mm. nor greater than 1.25 mm. Thickness of the bottom shall be not less than 0.75 mm. Tubes shall have a smooth, flat bottom, and shall stand vertically when placed on a flat, level surface. The open end of each tube shall be cut square, glazed, and finished without lip or constriction.

3.3 Thermal shock resistance.—When tested as specified in 4.3.2, the tubes shall not crack or break.

3.4 Workmanship.—Tubes shall be free of marked irregularities, striae, or bubbles that might interfere with their usefulness. All apparatus shall be annealed so that when tested as specified in 4.3.1, there shall be no ring strain. Any longitudinal strain, if present, shall be faint and highly diffused.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling. — Unless otherwise specified, at least twelve tubes shall be taken at random from each delivery for inspection and tests.

312307—55—No. 237

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

4.2 Inspection. — Tubes shall be inspected for material, style and design, and workmanship.

4.3 Tests.

4.3.1 *Strain examination.* — Tubes shall be tested in accordance with method No. 420 of Federal Specification DD–G–626 to determine compliance with requirements for strain.

4.3.2 *Thermal shock test.* — Tubes shall be tested in accordance with method No. 402 of Federal Specification DD–G–626 with a temperature differential of 79°C. The test shall be repeated five times.

5. PREPARATION FOR DELIVERY

5.1 Packaging.—Unless otherwise specified, test tubes shall be packaged in accordance with accepted commercial practice.

5.2 Packing.—The subject commodity shall be prepared for shipment to permit acceptance by carrier for transportation at the lowest applicable rate, and to afford maximum protection from normal hazards of transportation.

5.3 Marking.

5.3.1 *Packages.*—Unless otherwise specified, packages shall be marked with the name and size of the commodity, the quantity contained therein, and the name of the manufacturer and contractor.

5.3.2 *Shipping containers.*—Unless otherwise specified, shipping containers shall be marked with the name and size of the commodity as defined by the contract or order under which the shipment is made, the quantity contained therein, the name of the contractor, the number of the contract or order, the name and address of the consignee, and the gross weight.

6. NOTES

6.1 Ordering data.—Purchasers should exercise any desired option offered herein. (See 7.2.3.)

6.2 This specification covers the following item appearing in the Armed Services Medical Stock List:

6640–443–0300 Test Tube, Antigen Dilution, 17 mm. by 55 mm., Kahn, 12s.

6.3 It is believed that this specification adequately describes the characteristics to secure the desired material, and that normally no samples will be necessary prior to award to determine conformance to the requirements of this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 Federal specifications do not include all types, classes, grades, sizes, etc., of the commodities indicated by the titles of the specifications, or which are commercially available, but are intended to cover the type, etc., which are suitable for Federal Government requirements.

6.5 Transportation description.

6.5.1 *Transportation description applicable to this item is:*

Glassware, chemical laboratory, including test tubes

Carload minimum weight 18,000 pounds, subject to Rule 34, Uniform Freight Classification.

Truckload minimum weight 18,000 pounds.

Patent notice.—The furnishing by the Government of specifications, including drawings or other data in connection therewith, does not in any way grant the recipient of such material, or any other person, a license or right to manufacture, use, or sell any patented or unpatented invention that may in any way be related thereto.

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications and standard, of the issues in effect on date of invitation for bids, and special requirements, form a part of this specification for purchases made under this specification by the respective departments.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

7.2 Army, Navy, and Air Force.

7.2.1 *Applicable specifications and standard.*

Federal Specifications:

NN–B–621—Boxes, Wood, Nailed and Lock-Corner.

PPP–B–676—Boxes, Set-Up, Paperboard.

Military Specifications:

JAN–P–106—Packaging and Packing for Overseas Shipment—Boxes; Wood, Nailed.

JAN–P–120—Packaging and Packing for Overseas Shipment — Cartons, Folding, Paperboard.

MIL–L–10547—Liners, Case, Waterproof.

Military Standard:

MIL–STD–129—Marking for Shipment and Storage.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

7.2.2 *Packaging.*

7.2.2.1 *Unit of issue.* — Twelve (12) test tubes, as specified, constitute one unit.

7.2.2.2 *Unit packaging.* — Each unit (12 tubes) shall be packaged in a carton of appropriate size constructed in accordance with the requirements of Military Specification JAN–P–120 or Federal Specification PPP–B–676. Each carton shall be adequately secured to prevent accidental opening. Carton design shall include suitable interior fittings or sufficient cushioning material to insure against shifting or breakage. Each carton shall be adequately secured to prevent accidental opening.

7.2.4.1 *Unit cartons.*—Each unit carton shall be marked with stock number, nomenclature, unit, quantity, and name of contractor.

7.2.4.2 *Intermediate packages and shipping container.*—Each intermediate package and each shipping container shall be marked in accordance with Military Standard MIL–STD–129.

7.2.5 *Ordering data.* — Procurement documents should specify whether the subject commodity is to be packed for domestic or overseas shipment. (See 7.2.3.1 and 7.2.3.2.)

MILITARY INTERESTS:

Army—C M

Navy—Md

Air Force.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.