A-A-0051722(DM)
16 October 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

CATHETERIZATION PACK, URETHRAL, DISPOSABLE, WITH BOTTOM OUTLET DRAINAGE BAG

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a sterile, disposable, pre-assembled, ready-to-use urethral catheterization pack.

Salient characteristics:

General. The disposable catheterization pack shall be a complete set of items necessary for urethral catheterization, including a Foley type catheter and a bottom outlet drainage bag.

Characteristics. The catheterization pack shall consist of, as a minimum, the following disposable components:

Component	Quantity
Catheter, urethral (size as specified)	1
Drainage bag, with hanger, non-collapsible drip chamber, and permanently attached inlet and bottom outlet drainage tubes	1
Syringe, plastic, 10ml, pre-filled with a minimum of 9ml of sterile water	1
Examining gloves	1 pair
Underpad, absorbent	1
Fenestrated drape, absorbent	1
Prep pads, absorbent, or 12cm pre-moistened swab sticks	5
Device for attaching drainage tube to bedsheet	1
Jelly, lubricating, 5.0 grams (min)	1 pkg.
Solution, cleaning, Povidone-iodine, 10%, USP	22 cc (min)
Forceps, plastic 1/	1
Collection vial, urine, with cap	1
Label, urine collection vial	1
Tray, cleaning solution 2/	1
Instructions (excluding catheter insertion procedure)	1
List of contents (may appear on package)	1

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.

1/ Forceps need not be furnished if swab sticks are provided.
2/ The tray may be an integral part of the catheterization pack tray

The catheter shall be of the Foley design with a 5 cubic centimeter (cc) balloon capacity. The catheter shall meet all requirements of American Society for Testing and Materials (ASTM) Standard F623, Standard Performance Specification for Foley Catheter and shall be tested for physical properties specified herein. The catheter shall be provided in the following sizes, as specified:

Size	Outside diameter (French)
1	16
2	18

The catheter shall be permanently and legibly marked with the manufacturer's name or registered trade-mark, French size, and balloon capacity.

The disposable plastic syringe shall be pre-filled with a minimum of 9 cc of sterile water. The syringe shall be marked as follows: "CAUTION: Sterile Water-for Balloon Inflation Only," or words to this effect. The syringe shall be marked in 0.2 cc graduations, with each 1.0 cc graduation numbered.

The urine collection bag assembly shall consist of the urine collection bag, inlet tubing, non-collapsible drip chamber, anti-reflux valve, and bottom drainage tube. The inlet tubing shall be permanently attached to the drip chamber. The inlet tubing shall be 0.25 inch (minimum) inside diameter and shall have a tapered connector for attaching to the catheter. The anti-reflux valve shall be positioned at the top of, or integral to, the drip chamber. The drainage tube shall be manufactured from rubber and shall be permanently attached to the bottom of the collection bag. The bottom drainage tube shall have a shut-off clamp affixed to it. The clamp shall be capable of one-handed operation, and shall not cut the tube. The bag shall have a pocket located directly above the bottom drainage tube for insertion of the tube tip.

The urine collection bag shall be vented with a filter which prevents passage of liquid out of the bag. The front of the collection bag, the inlet tubing, and the drip chamber shall be sufficiently transparent or translucent to observe the flow of the contents. The bag shall be permanently marked with the manufacturer's name or registered trademark. The bag shall be marked in 100 cc increments to a minimum of 2000 cc, but not exceeding 2500 cc. A minimum of every other graduation shall be

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.

identified. In addition, the lower left corner of the bag shall be marked in 25 cc increments to a minimum of 100 cc, with a minimum of every other graduation identified. The scale in the lower left corner shall be made 45 degrees to the bottom of the bag. A device for securing the inlet tubing to bed sheet shall be attached to the tubing with a rubber band.

The examining gloves shall be of universal size, pre-powdered, and folded or packaged to facilitate donning without violating asceptic technique. The gloves shall be free of seams or welts in the palm or finger tip areas.

The absorbent underpad shall be coated on one side with a impermeable plastic film. The underpad shall be 48 cm by 62 cm (minimum).

The absorbent drape shall have a fenestrated center, and shall be 33 cm by 46 cm (minimum).

The prep pads shall be absorbent balls 4 cm (minimum) in diameter. As an alternative, pre-moistened 12 cm swab sticks may be supplied in a suitable wrapping to retain moistness.

The lubricating jelly shall be sterile, bacteriostatic, and water-soluble. A minimum of 5 grams of lubricating jelly shall be provided in a suitable foil packet.

Material. The catheter shall be manufactured from durable, soft latex rubber, natural or cream color. The syringe, tubing, and drip chamber shall be manufactured from a suitable plastic. The examining gloves shall be manufactured from an elastomeric film or sheet plastic. The absorbent drape and underpad shall be manufactured from facial tissue or other suitable non-woven material and shall be multiple-layer. The prep pads or swabs shall be manufactured from cotton or viscose rayon.

Workmanship. All seams and joints shall be leakproof. There shall be no defects that will affect the life, serviceability, or appearance of any component of the catheterization pack.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One catheterization pack packaged sterile in a sealed peel-open container, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Biocompatibility Test. A certificate of compliance, stipulating that offered items meet the biocompatibility testof the ASTM Standard F623, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date of actual testing, the test method and results, the name and title of the person evaluating results, the current contract number and NSN, a statement to the effect that material processing and formulation have not changed since actual testing was performed, and the signature and title of the person making the certification.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the currentversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one catheterization pack.

Catheter. Catheters shall be tested in accordance with the performance tests of ASTM F623.

Catheters shall be tested for breaking strength and ultimate elongation in accordance with paragraphs 1 through 4 of the applicable Method of Federal Test Method Standard 601. Minimum test values and test methods shall be as follows:

	Test Method	Size 1	Size 2
*breaking strength (pounds)	4111	102	127
ultimate elongation (percent)	4121	750	750

*same as breaking force

Two specimens shall be cut from each catheter. One specimen shall be tested as cut and one specimen shall be subjected to accelerated aging. The test value obtained from each specimen shall constitute one unit test result.

The catheters shall be subjected to accelerated aging in accordance with Test Method 7221. The breaking strength and ultimate elongation shall again be measured. The test values shall not decrease by more than 25 percent, nor shall the rubber crack or become brittle or tacky. Sampling shall be S-1. Acceptance for unaged physical properties testing shall be based upon the arithmetic average of the values obtained. The change in physical properties between the aged specimen and unaged specimen cut from the same catheter shall be calculated. The AQL for aged versus unaged properties testing shall be 1.0.

Urine collection bag assembly. The same apparatus used for the Flow Rate Test of the Foley Catheter in ASTM F623 shall be applied to the urine collection bag assembly as follows. The catheter shall be cut between the funnel and balloon. The funnel shall be attached to the tubing connector. The other end of the catheter shall be connected to the on/off valve of the reservoir. The collection bag shall be attached upright to the bench surface with the safety chamber directly under the reservoir outlet on/off valve. The vertical distance from the reservoir outlet to the safety chamber shall be 40 cm $\pm$ 5 cm. Water flow into the reservoir shall be such that the height of the water in the reservoir remains a constant 20cm. The excess inlet tubing shall be laid out horizontally on the bench top.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The on/off valve shall be opened. The bag shall collect 2000cc of water in less than 21 minutes. The tubing shall then be disconnected from the on/off valve and the collection bag inverted so that the bottom drainage tube is on top. The bag shall remain in this position for 5 minutes. There shall be no leakage past the anti-reflux valve or out of the bag vent. There shall be no leakage from any joint or seam of the assembly during the course of this test. Sampling shall be S-1 with an AQL of 0.25.

Catheter adapter. The catheter shall be attached to the catheter adapter. The catheter shall be subjected to an 8 pound tensile force for 1 minute. The catheter shall not separate from the adapter. Five catheter/catheter adapters shall be tested from each lot, with no defects permitted.

Sterility. All of the components of the catheterization pack shall be sterile when tested in accordance with the USP.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-149-0104	CATHETERIZATION PACK, URETHRAL, Disposable, With Bottom Outlet Drainage Bag, 16 French Catheter
6515-00-149-0105	CATHETERIZATION PACK, URETHRAL, Disposable, With Bottom Outlet Drainage Bag, 18 French Catheter

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 05121

Project No. 6515-3465

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 704-036/45824

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.