A-A-0051736 (DM)
3 November 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

DENTAL ENGINE, BENCH TYPE, HIGH SPEED, 115 Volt, 60 Hz, AC

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by DLA, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers an electrical bench engine suitable for use in the dental prosthetic laboratory during the fabrication of dental prosthesis.

Salient characteristics:

The bench engine shall be Teledyne Hanau's Emesco Bench Engine Model No. 92NP with Emesco No. 9 Foot Controller or equal, consisting essentially of a bench mounting base, motor assembly, short triple section belt arm, foot activated rheostat for control, suitable electrical connectors and applicable accessories. The weight and weight distribution of the engine (including bench mounting base) shall be such to provide proper balance and preclude tilting or moving of the engine during use. The engine with foot controller, shall be suitable for use on 110 volt, 50/60 hertz, AC electrical power. The motor shall be an AC/DC motor which shall have a minimum rating of 1/5 horsepower and shall be provided with a 2-step drive pulley. The engine shall be reversible and be capable of driving a dental handpiece from 1,200 to 34,000 rpm. Belt arms shall be triple arm, all cord type, so constructed as to compensate tension on belt. The triple arm shall consist of a 12 inch long main arm which shall be the arm at the motor end, a 10-1/2 inch long upper arm which shall be the center arm, and a 5-1/2 inch long lower arm which shall be the arm at the handpiece end. The triple arm shall have six ball bearing pulleys. Belt guards shall be provided for all pulleys. The foot activated speed control mechanism shall be a solid state silicon controlled rectifier type rheostat enclosed in a housing 6-1/2 inches in diameter and 2-1/4 inches high and shall weigh 5-1/2 $\pm$ 1/2 pounds.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

A tolerance of ± 1/2 inch shall be applicable to all specified dimensions herein. The electrical supply cord shall be no less than 8 feet long, rubber-covered 3 wire cord, with a 3-prong plug suitable for use with receptacles specified in American National Standards Institute (ANSI) Standard C73.11 for caps and receptacles, 125 volts, 15 amperes, 2 pole, 3 wire, grounding type. Bench engine shall conform to the requirements of Underwriters Laboratories, Inc. standards. Bench engine base and motor housing shall have a bright chrome finish. All other components shall be of corrosion resisting steel, or metal suitably plated, painted, or treated to provide corrosion resistance. The engine shall be provided complete with one engine belt (6 feet 4 inches). Each unit shall be provided with service data which shall include installation, operation, maintenance and repair, and parts list for the motor, controller, and arm including schematic diagrams.

Workmanship. Dental engines shall be free from defects which detract from their appearance or may impair their serviceability.

Unit. Each (EA). One bench type, high speed dental engine, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one bench engine. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-00-484-9494	DENTAL ENGINE, BENCH TYPE, HIGH SPEED 115 Volt, 60 Hz, AC

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordinating Activities:
VA-OSS
PHS

47779

Preparing Activity:

DoD - MB

Agent:

DLA - DM

Project No. 6520-1983

Location: NCR Disk #2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide sugg ons. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and a In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-,RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-0051736 (DM)	DENTAL ENGINE, BENCH TYPE, HIGH SPEED, 115 Volt, 60 Hz, AC

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER (Specify) _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -> Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.