METRIC
A-A-51707A
31 August 1993
SUPERSEDING
A-A-51707
8 October 1986

COMMERCIAL ITEM DESCRIPTION

LENS, OPHTHALMIC, PLASTIC

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers the plastic lenses made primarily of allyl diglycol carbonate monomer (ADC), commonly called CR-39 hard resin lenses. These lenses are of dress type for visual deficiency correction.

Salient characteristics:

Type:

I. Lens blank semi-finished
II. Lens, finished, uncut

Class:

A. Bifocal
B. Single vision

Style:

1. Spheres
2. Plano minus cylinders
3. Plus or minus compounds
4. Minus or minus compounds
5. 28 mm flat top
6. 35 mm flat top

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The lens diameter under this specification shall be 71 mm ± 1 mm diameter for the single vision and the bifocal, except as noted in the specification. Other shapes are acceptable, provided equivalent size is met. All lenses shall be clear white, free of any visible color tint.

Refractive power tolerance for the finished lens shall be as follows:

Sphere	0.00 thru 6.50 diopter is ± 0.13D Greater than + 6.50 diopter or less than - 6.50 diopter is ± 2%
Cylinder	-0.25 thru -2.00 diopter is ± 0.13D -2.12 thru -4.50 diopter is ± 0.15D less than -4.50 diopter is ± 4%

All plastic lenses shall be capable of passing the impact resistance specified in the current edition of ANSI Z80.1, Recommendations for Prescription Opthalmic Lenses. For Type II finished lenses, contractor shall, when requested, pull sample(s) from each lot and demonstrate the impact resistance test for the contracting officer's representative. For Type I semi-finished lenses, the contractor shall, when requested, allow the contracting officer's representative to inspect quality assurance records demonstrating the capability of surfaced Type I lenses to pass the impact resistant test.

All plastic lenses shall be clear white made primarily of Allyl Diglycol Carbonate (ADC). The lens material shall contain not less than 85% ADC and have an index of refraction of 1.498 nominal and a specific gravity of 1.32 nominal. Lenses shall be free of bubbles, inclusions, striae, and cloudiness, except that small edge bubbles which do not adversely effect lens utilization are allowed. Lenses shall remain clear and free of discoloration under normal use (including outdoor wear) for a period of one year from date of shipment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Type II, Class B single vision lenses shall be a member of an integrated series of minus-toric opthalmic lenses having surface curvatures computed to conform to the corrected-curve lens series. Minimum center thickness shall be 1.8 mm and minimum edge thickness shall be 1.0 mm measured 1-1/2 mm from the edge of lens; the edge thickness measurement is taken parallel to the lens axis.

Type I, Class A lens blanks shall be finished on outside of the lens with the segment size of 28 mm wide x 19 mm high with its straight edge positioned 6.0 mm ± 0.5 horizontal midline and 6.0 mm ± 1.0 inward from the vertical midline of the carrier lens for the 28 mm flat top lens blank. As for the 35 mm flat top lens blank, the segment size shall be 35 mm wide x 22 mm high with its straight edge positioned 6.0 mm ± 0.5 horizontal midline and 6 mm (+ 0.5 or -1.0) inward from the vertical midline of the carrier lens.

Unless otherwise approved by the contracting officer, finished lens shall be with the front base curve specified on the individual lens description, tolerance for the front base curve is ± 1.0 diopter.

Workmanship. The lens shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Pair (PR). Two lenses (1 pair), as specified, constitutes one unit. Each lens shall be individually packaged in a snug fitting envelope.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one lens as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one lens. Sampling shall be inspection level S-1 with an AQL of 1.5.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.
The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	Type	Class
6540-00-001-0727 (lead)	Lens, Ophthalmic, Plastic	II	B
6540-00-001-0215 (lead)	Lens, Ophthalmic, Plastic	II	B
6540-01-375-4090 (lead)	Lens, Ophthalmic, Plastic	I	B
6540-01-375-4058 (lead)	Lens, Ophthalmic, Plastic	I	A
6540-01-375-0095 (lead)	Lens, Ophthalmic, Plastic	I	A
6540-00-001-0193 (lead)	Lens, Ophthalmic, Plastic	II	B
6540-00-001-0100 (lead)	Lens, Ophthalmic, Plastic	II	B

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6540-1264
Location: Intergraph/lens

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

   The submitter of this form must complete blocks 4, 5, 6, and 7.

3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51707A	930831

3. DOCUMENT TITLE
LENS, OPHTHALMIC, PLASTIC

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.