A-A-51709
31 October 1986

COMMERCIAL ITEM DESCRIPTION

THERMOMETER, CLINICAL, HUMAN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers the requirements for maximum self-registering, mercury-in-glass reusable thermometers suitable for body temperature measurements in humans.

Salient characteristics:

Scope. Thermometer covered by this document shall be of the following types and styles:

Type I - Oral
Type II - Rectal

Style 1 - with case
Style 2 - without case

Requirements. The thermometer shall be of either oral or rectal type with a stubby bulb design, as specified, suitable for human body temperature measurements in Fahrenheit scale. The stem shall be triangular in cross section with magnifying lens at the junction of two sides with the numerals and graduations placed on the sides adjacent to the magnifying lens.

Except otherwise provided in this document, the thermometer shall conform to all applicable requirements of ASTM E667 for Clinical Thermometers (Maximum self-registering, mercury-in-glass).

In addition to the requirements of the ASTM Standard, the following shall apply.

Stem. The stem shall be of lens front thermometer tubing having white enameled backing. Thermometer intended for oral use shall be identified by a distinct blue glass bead fused to the top of the stem. Thermometer intended for rectal use shall be identified by a distinct red glass bead fused to the top of the stem.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Mercury. All thermometers in which there is sufficient gas or other foreign material to cause separation of the mercury column above the constriction during shipment, test, or use, shall be classified as defective.

Markings. All markings shall be permanent and legible and shall be applied without etching, by application of a stain, enamel or other suitable material fired into the glass. The graduation marks shall be substantially straight, regularly spaced, of uniform width, and perpendicular to the axis of the stem. When tested as specified in 6.2 of ASTM E667, the markings shall not fade or discolor in such a manner as to impair their legibility. The loss of one complete graduation or numeral shall be considered a cause for rejection.

Case. A case shall be supplied with each style 1 thermometer. Case shall consist of a barrel and a cap, constructed of aluminum, nickel or chromium plated metal hard rubber or plastic, and shall contain suitable interior fittings or cushioning material to individually protect each thermometer from damage. Overall case length shall not exceed 4-7/8 inches. Case shall be free of deformity and sharp edges.

Order of precedence. In the event of a conflict between the requirements of the ASTM E667 and other requirements described herein, the requirements specified in this document shall take precedence.

Workmanship. The thermometer shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests or obtaining certificates of conformance from his suppliers to assure that materials are in accordance with requirements specified

Unit.

Type I, Style 2 and Type II, Style 2. Package (PG). One package containing six thermometers of one type only, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the currentversion before use.

Type I, Style 1 and Type II, Style 1. Each (EA). One thermometer with case, as specified, constitutes one unit.

Type I, Style 2 and Type II, Style 2. Each (EA). One thermometer without case, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one thermometer. Inspection level shall be S-2 with an AQL of 1.5.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:00Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-149-1404	THERMOMETER, HUMAN, RECTAL, 6's, Type II, Style 2
6515-00-149-1405	THERMOMETER, HUMAN, ORAL, Each, Type I, Style 1
6515-00-149-1406	THERMOMETER, HUMAN, ORAL, 6s, Type I, Style 2
6515-00-149-1407	THERMOMETER, HUMAN, RECTAL, Each, Type II, Style 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

In addition, this document also covers the requirements for the following Veterans Administration (VA) items:

NSN	ITEM IDENTIFICATION
6515-00-226-6985	THERMOMETER, HUMAN, ORAL, Type I, Style 2
6515-00-990-1971	THERMOMETER, HUMAN, RECTAL, Type II, Style 2

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MQPAS

Project No. 6515-3760

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51709	THERMOMETER, CLINICAL, HUMAN

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.