A-A-51711
16 October 1986

COMMERCIAL ITEM DESCRIPTION

VISCOMETER, CAPILLARY TYPE, Cannon-Fenske Reverse Flow

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers Cannon-Fenske, reverse flow, capillary type viscometers, calibrated, for use in determining viscosities of opaque liquids.

Salient characteristics:

The viscometer shall be in accordance with the applicable portions of ASTM designation D-446.

Shall be constructed of fully annealed, low expansion borosilicate glass, free of cracks, chipping, protrusions, stones, bubbles, seeds or striae.

Each viscometer shall be marked with size-number, serial-number, and manufacturers designation.

Etch all timing marks and fill with an opaque color, or otherwise make a permanent part of the viscometer.

The viscometer shall be designed to fit through a 51 millimeter (mm) diameter hole in the lid of the constant temperature bath.

Each viscometer shall be calibrated in accordance with ASTM designation D-446, and shall be supplied with a Calibration Certificate and instructions for use. Calibration shall be by comparison with a master standard. Master standard shall be based on a viscosity of 1.0038 centistokes for water at 68 degrees Fahrenheit (20 degrees Centigrade) and shall be calibrated in accordance with ASTM designation D-2162. Accuracy of the master standard shall be traceable to National Bureau of Standards requirements. Accuracy of the viscometer shall be as indicated by calibration $\pm$ 0.2 percent. Test results and calibration results may be averaged for calibration certificate, providing test results are within specification requirements.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Reference specification. Where ASTM designations are referenced herein, the issue of that designation on the date of solicitation shall be used to determine compliance with stated requirements.

Workmanship. The viscometer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One viscometer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The viscometer shall be examined to determine compliance with all requirements contained in this specification.

Testing. Testing shall be conducted to determine compliance with specification requirements. Shall be in accordance with the applicable portions of ASTM Methods of Test D445 and D446.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one viscometer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6630-00-359-9887	Size 50
6630-00-359-9888	Size 100
6630-00-359-9889	Size 200
6630-00-359-9891	Size 400
6630-00-397-9647	Size 150
6630-00-398-3736	Size 350
6630-00-398-5118	Size 450

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6630-0442
Location: CID/VISCOMETER, CAPILLARY TYPE/NCR

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.