A-A-51723
28 November 1986

COMMERCIAL ITEM DESCRIPTION

METER FOOTCANDLE, PHOTOELECTRIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a photoelectric footcandle meter suitable for the measurement of light.

Salient characteristics:

The meter shall be modularized, solid state, and portable.

The meter shall measure collimator light beam illuminance with state-of-art sensitivity, resolution, and accuracy.

The meter shall have no controls required to operate throughout specific performance ranges.

The meter shall have overall measurements that do not exceed 6-1/2" high, 3-1/2" wide, and 2" deep, and weigh no more than 2 pounds.

The meter shall operate from internal battery power supply. It shall be supplied without battery.

Unit shall measure illuminance from 0 to 50 footcandles minimum.

Unit accuracy shall be 5 percent of reading or better.

Unit shall be digital readout.

Subassemblies and chassis components shall be removable without unsoldering any other hard-wire subassembly or printed circuit card or component.

AMSC N/A FSC 6695

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The item shall perform within the following climatic conditions:

a. Temperature, non-operating; -46°C to +70°C.

b. Temperature, operating; 0°C to 55°C.

c. In an environment of 95 percent humidity.

d. In an environment of 30 mppcf or 10 mg/m3 of total dust less than 1 percent quartz or 5mg/m3 respirable dust.

Unit shall be furnished with a protective case.

Chemical agent resistant coating is not required.

The meter shall be fully cosine corrected 0-90 degrees.

Workmanship. The photoelectric footcandle meter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One meter, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6695-01-217-7647

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC

Project No. 6695-0070

Location: CID/METER/H17-5

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.