A-A-51726
31 October 1986

COMMERCIAL ITEM DESCRIPTION

TEST KIT, OCCULT BLOOD AND GASTRIC pH DETERMINATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a test kit for performing visual tests for monitoring gastric pH in conjunction with determining the presence of occult blood in a sample of gastric aspirate or vomitus.

Salient characteristics:

Shall be a test kit for determining, by use of a chemically treated filter paper, the presence of occult blood and the pH of gastric aspirate or vomitus.

The contents of the test kit shall be sufficient for conducting forty (40) determinations (tests), and shall include the following:

1. Cardboard slides with filter paper that is treated with guaiac resin (occult blood test area) and pH indicator (pH test area).

2. A dispenser bottle of ready-to-use developer solution containing 2.5% hydrogen peroxide and 30% denatured alcohol in a citrate-buffered aqueous solution.

3. Specimen applicators.

4. A package insert with instruction and color guides.

The filter paper on the slide shall include:

1. An occult blood test area.

2. Positive and negative controls to check the sensitivity of the occult blood test area and the effectiveness of the developer solution.

AMSC/NA FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

3. A pH test area.

4. A pH color chart that includes color standards for pH values of 1, 2, 3, 4, 5-7 and over 7.

The occult blood test area on the filter paper shall react in accordance with the following, when a sample of specimen (gastric aspirate or vomitus) and developer are applied to the area in accordance with manufacturer's instruction:

1. A color reaction within 60 seconds occurs if the sample contains blood.

2. No color reaction when there is no blood in the sample.

The pH test area on the filter paper shall produce a color reaction when a sample of specimen is applied to the area in accordance with manufacturer's instructions. The resultant color shall be indicative of the pH of the sample. The pH of the sample shall be determined by matching the color with the pH color chart on the test slide.

The tests, conducted on the filter paper shall be unaffected by low pH, normal therapeutic concentrations of cimetidine, iron or copper salts.

Workmanship. The test kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One test kit, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspection and testing shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspection and testing. Sampling for inspection and testing shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following products as of the date of this document, are purported by the following companies to meet the Salient characteristics of the item described herein:

Smith-Kline Diagnostics Product No. 66000

NOTE: The following National Stock Number is covered by this document:

6550-01-163-6334

Ordering data. (Expiration dating required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6550-0055

Location: CID/TK, OCC.BLO
NCR disk, 7

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.