A-A-51727
3 November 1986

COMMERCIAL ITEM DESCRIPTION

SEALING COMPOUND, PULP CANAL, DENTAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a temporary filling material, suitable for use in sealing access openings to root canals.

Salient characteristics:

Sealing compound shall consist of a mixture of zinc oxide, calcium sulfate, zinc sulfate, glycol-acetate, polyvinyl acetate, polyvinyl chloride acetate, triethanolamine, and red pigment, in paste form. The sealing compound shall be a non-toxic, ready-mixed temporary filling material, suitable for use in sealing access openings to root canals between treatments. The sealing compound shall be light pink in color.

pH. The pH shall be 6.90 ± .05 when tested in accordance with ASTM E 70, and prepared as specified in test section under pH.

Linear setting expansion. The linear setting expansion shall be 14.2% ± 0.1%, when tested as specified in test section under linear setting expansion.

Working properties.

The sealing compound shall have a plastic consistency. Consistency shall be such that the compound shall stay in place when inserted in the root canal opening.

The sealing compound shall harden and expand upon contact with saliva or water. Rate and degree of hardening shall be such that the patient can chew on the filling 1 hour after it has been inserted. Expansion shall be such that the filling provides a leakproof seal for the access opening to the root canal.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Performance certification. The offeror shall furnish a statement to the procuring activity, prior to award of contract or purchase order, that the sealing compound has been clinically tested and found effective in providing a chewing surface in sealing access openings to root canals and is free of toxicity as required. Supporting data shall be forwarded with the statement or made available on request by the procuring activity. If the data furnished is found inadequate, additional information shall be furnished on request. The Government reserves the right to examine the data, to consult with the test activity, and to determine whether such data is satisfactory and acceptable.

Instruction. Adequate and accurate instructions for use shall be included with each package. Instructions shall include the information that the patient can chew on filling 1 hour after treatment.

Workmanship. Sealing compound shall show no separation, shall be free of chemical or color changes, hardening in the tube, or any other defect which detracts from its appearance or may impair its serviceability.

Unit. Package (PG). One package containing four tubes, each tube containing a minimum of 7 grams of sealing compound and a key which is used to facilitate dispensing of sealing compound, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one package of four tubes, each tube containing a minimum of 7 grams each. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

pH. Spatulate 1 gram of sealing compound with sufficient water to induce a rapid set. Crush the set compound into fine particles and stir vigorously into 10 milliliters of distilled water. Allow particles to settle and determine pH of clear supernatant liquid within 24 hours.

Linear setting expansion. Twelve cylindrical specimens of the sealing compound measuring 6 mm x 12 mm are placed in open plastic tubes and immersed in a distilled water bath for ten days at 37° ± 1° Centigrade. On day eleven, measure length and determine the percent of linear expansion. The largest and smallest length measurement shall be dropped. The average of the remaining 10 shall be used to calculate the linear setting expansion. The length measurement shall be taken to the nearest 0.5 mm.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

Stock Number	Item Identification
6520-01-060-9516	SEALING COMPOUND, PULP CANAL, DENTAL, Paste, 4s

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-1946

Location: NCR Disk 2

49044

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.