A-A-0051734(DM)
24 October 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

WRAPPER, STERILIZATION

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers two sizes of heavy duty, disposable, polypropylene sterilization wrappers.

Size 1 - 45 x 45 inches (114 x 114 cm)
Size 2 - 54 x 54 inches (137 x 137 cm)

Salient characteristics:

The sterilization wrappers shall be clean, disposable polypropylene, heavy duty, consisting of two sizes, fabricated of a nonwoven material, suitable for wrapping and processing surgical supplies and instrument sets for sterilization. The wrappers shall be 2.1 oz./sq. yard.

The wrappers shall be fluid repellent, memory-free, low linting, puncture and abrasion resistant, and hypoallergenic. They shall have bacterial barrier properties.

The wrappers shall be freely permeable to steam and ethylene oxide penetration. The wrappers shall assure sufficient concentration of steam to guarantee sterilization of the contents when used as a double wrapper on a surgical pack. Adequate air removal for drying efficacy shall be insured.

There shall not be more than a three percent weight gain during steam sterilization of a surgical pack for drying efficacy.

The pack density standard shall be no longer than 7.2 lbs./cubic foot.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The wrappers shall maintain the supplies in a sterile manner for thirty days without a "dust cover" in a clean, temperature-controlled, humidity-controlled environment if the wrapper has not been opened or tampered with.

The wrappers shall have a surface resistivity no greater than 1 x 10^{11} ohms/in.^{2}.

The wrappers shall meet conductivity standards NFPA 99, and flame resistance standards NFPA 702, Class I.

Workmanship. The sterilization wrappers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 100 wrappers of size one or 50 wrappers of size two, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one (1) package.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-153-6437
6530-01-084-1456

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

YQP 70009

Project No. 6530-1848

Location: CID/WRAPPER/H1A-5

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.