INCH-POUND
A-A-51735A
17 May 1992
SUPERSEDING
A-A-51735
14 November 1986

COMMERCIAL ITEM DESCRIPTION

RESUSCITATOR, HAND-OPERATED

(Intermittent Positive Pressure,
High Oxygen Concentration Capacity)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a hand powered pulmonary resuscitator intended for use during emergency situations.

Salient Characteristics:

The resuscitator shall be in conformance with ASTM F 920 for adult resuscitators and shall not include a pressure limiting system. It shall include the following items:

* One (1) adult compression bag of the folding type for storage.

* One (1) oxygen reservoir bag assembly, closed end type. Open end accumulator type shall not be acceptable.

* One (1) patient valve with an approximate 90 degree angle between inlet and outlet. The outlet shall be a 360 degree male swivel patient connector in accordance with ASTM F 920 and ASTM F 1054.

* Two (2) face masks, reusable; one small adult size, and one medium adult size (single patient use type masks shall not be acceptable). The face masks shall be in accordance with ASTM F 920 and ASTM F 1054.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

* Three (3) translucent oral pharyngeal airways; one size 70 mm, one size 80 mm, and one size 100 mm. The airways shall be the "Guedel" type and have the center open to accommodate the insertion of a suction catheter.

* One (1) lightweight, semirigid plastic carrying case with handle and hinged lid with a snap-action or hasp type latch. Dimensions of the case shall not exceed 6 inches by 6.5 inches by 13 inches in order to fit into field medical assemblages.

* One (1) operating, service and maintenance manual.

All components in contact with the patient breathing mixture shall be re-usable and capable of being cleaned and sterilized/disinfected in accordance with ASTM F 920 and by a method defined in the U.S. Pharmacopeia - National Formulary. Single patient use or disposable components shall not be acceptable.

The compression bag shall be translucent and allow the operator to monitor any collection of patient fluid or otherwise monitor results of improper valving action.

Hard parts of housings for all valves shall be fabricated of a suitable transparent plastic material allowing the operator to monitor valve action.

In addition to ASTM F 920 requirements for storage, the resuscitator with components specified herein shall have a storage life of at least 5 years without degradation of performance. Elastomeric parts, inadvertently (re)packaged having a deformed shape, shall return to their intended shape when unpacked without exhibiting memory characteristics.

In addition to ASTM F 920 requirements for disassembly and reassembly, all components intended to be disassembled by the operator shall be sized/shaped to prohibit incorrect assembly. Tools shall not be required for this disassembly or assembly.

Workmanship. The resuscitator shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One resuscitator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for examination purposes shall be one resuscitator.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for test purposes shall be one resuscitator.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include all applicable tests in ASTM F 920.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-204-5394

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5201

Location: RESU12B1.CID

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.