A-A-51743A
26 February 1988
SUPERSEDING
A-A-0051743(DM)
4 October 1986

COMMERCIAL ITEM DESCRIPTION

CANISTER ASSEMBLY, SUCTION, SURGICAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable suction canister with locking lid suitable for use during and after surgical procedures.

Salient characteristics:

The disposable suction canister with locking lid shall be suitable for use during and after surgical procedures. The canister shall be fabricated of a high impact, transparent plastic material. It shall have a minimum capacity of 2000 ml. The canister shall have graduation markings at least every 100 ml from 100 ml to 2000 ml. It shall be provided with a frosted area or other suitable means for recording patient data.

The lid shall be fabricated of a high impact plastic material. It shall be provided with two tapered ports designed to accept 8 mm I.D. tubing. One port shall be marked "VACUUM" and the other "PATIENT". The vacuum port shall have a shut-off valve to prevent overflow.

The lid and canister shall be capable of tandem hook-up. The canister shall have a bacterial filter capable of removing at least 99 percent of aerosol particles while not shutting off the suction, if the filter becomes saturated. The canister shall be treated to prevent static electricity build-up.

Workmanship. The canister shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 40 canisters with lids, as specified, constitutes one unit.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

First article inspection. First article inspection shall be performed on the canister assembly when a first article sample is required. This inspection shall include the examination and tests.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 , with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one canister assembly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Defects.

Any component part missing, broken or otherwise not usable.

Exposed edge not free or burrs or sharpness.

Unit does not conform to salient characteristics.

Identification marking not complete, permanent, and legible.

Finish not free of chips, cracks, scratches and other appearance defects.

Sampling for tests. Sampling for tests including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one canister assembly.

For tests. Sampling for tests shall be conducted as follows:

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a first article sample has successfully passed all tests, performance test shall not be required at acceptance.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-152-5512

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4020

PGC 46983

Location: CID/CANISTER/H1C-12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.