Fed. Std. No. 204a
September 4, 1964

(GG-S-764, Issue in Effect)
(MIL-W-6510, Issue in Effect)
SUPERSEDING
Int. Fed. Std. No .00204 (GSA-FSS)
April 4, 1963

FEDERAL STANDARD

STOP WATCH, LABORATORY

Authority. *This standard is issued pursuant to the Federal Property and Administrative Services Act of* 1949, *as amended, and its application to the purchase of commodities referred to herein is mandatory on all Federal agencies.*

S1. Purpose and scope. This standard establishes standard types of laboratory stop watches and adopts them as standard items for use by all Federal agencies to achieve highest practical standardization.

S2. Application. Where the items in this standard will meet the requirements of using agencies, purchases from commercial sources under Federal Specification GG-S-764 and Military Specification MIL-W-6510 should be limited to items which have characteristics conforming to the standard items specified herein. Existing Government stocks of items with other characteristics should continue to be utilized until exhausted. In those cases where this Federal Standard eliminates supply items and the items so eliminated are under contract, it is recommended that the contract be negotiated or amended to conform to this standard.

S3. Identification of standard items. Applicable Federal Stock Numbers and descriptions of items covered by this standard are as follows:

FSN 6645-243-9185—0.1 Sec. to 60 Sec. dial graduation, 15 min. register, 16 size, 7 jewel, noncontinuous running type, 1 sweep hand, white dial, black numerals and graduations, single action, push button and side slide control, plastic crystal; Federal Specification GG-S-764, type I, class 2.

FSN 6645-244-0359—0.2 sec. to 60 sec. dial graduation, 30 min. register, 16 size, 7 jewel, noncontinuous running type, 1 sweep hand, white dial, black numerals and graduations single action, stem push button control, plastic crystal; Federal Specification GG-S-764, type I, class 2.

FSN 6645-250-4680—0.2 sec. to 60 sec. dial graduation, 30 min. register, 16 size, 7 jewel, continuous running type, 1 sweep hand, white dial, black numerals and graduations, single action, stem push button control, plastic crystal; Federal Specification GG-S-764, type I, class 2.

FSN 6645-551-4001—0.6 sec. to 60 sec. dial graduation, 30 min. register, 16 size, 7 jewel, noncontinuous running type, 1 sweep hand, white dial, black numerals and graduations, single action, stem push button and side slide control, plastic crystal; Federal Specification GG-S-764, type I, class 1.

FSN 6645-577-0322—0.1 sec. to 10 sec. dial graduations, 10 min. register, 16 size, 11 jewel, noncontinuous running type, 1 sweep hand, black dial, white numerals and graduations, stem push button control; Military Specification MIL-W-6510, type A-8.

FSN 6645-899-1235—0.2 sec. to 60 sec. dial graduations, 30 min. register, minimum,

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:55Z
Check the source to verify that this is the current version before use.

16 size, 7 jewels minimum, continuous running type, 1 sweep hand, single action, white dial, crown or push-piece control, colorless transparent crystal; Federal Specification GG-S-764, type I, class 3.

S4. Changes. When a Federal agency considers that a Federal standard does not provide for its essential needs, written request for adding to or otherwise changing the standard, supported by adequate justification, shall be sent to the Administration. This justification shall explain wherein the standard does not provide for essential needs. The request shall be sent in duplicate to the General Services Administration, Federal Supply Service, Standardization Division, Washington D. C. 20407. The Administration will determine the appropriate action to be taken and will notify the agency.

S5. Conflict with referenced specifications. Where the requirements specified in this standard conflict with any requirement in a referenced specification, the requirements of the standard shall apply. Nature of conflict between the standard and the referenced specification shall be submitted, in duplicate, to General Services Administration, Federal Supply Service, Standardization Division, Washington, D.C. 20407.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:55Z
Check the source to verify that this is the current version before use.